University of Dundee

A study of the effect of clinical washing decontamination process on corrosion resistance of Martensitic Stainless Steel 420

Xu, Yunwei; Huang, Zhihong; Corner, George

Published in:
Bio-medical Materials and Engineering

DOI:
[10.3233/BME-161590](https://doi.org/10.3233/BME-161590)

Publication date:
2016

Document Version
Peer reviewed version

[Link to publication in Discovery Research Portal](#)

Citation for published version (APA):

Xu, Y., Huang, Z., & Corner, G. (2016). A study of the effect of clinical washing decontamination process on corrosion resistance of Martensitic Stainless Steel 420. *Bio-medical Materials and Engineering*, 27(4), 341-351. <https://doi.org/10.3233/BME-161590>

General rights

Copyright and moral rights for the publications made accessible in Discovery Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from Discovery Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain.
- You may freely distribute the URL identifying the publication in the public portal.

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

**A Study of the Effect of Clinical Washing Decontamination
Process on Corrosion Resistance of Martensitic Stainless
Steel 420**

Yunwei Xu, Zhihong Huang, George Corner

School of Science and Engineering, University of Dundee,
Dundee, UK

Corresponding Author

Yunwei Xu

Tel: +44 7450242877

Email: ywzxu@dundee.ac.uk

ABSTRACT:

Corrosion of surgical instruments provides a seat for contamination and prevents proper sterilisation, placing both patients and medical staff at risk of infection. Corrosion can also compromise the structural integrity of instruments and lead to mechanical failure in use. It is essential to understand the various factors affecting corrosion resistance of surgical instruments and how it can be minimised.

This paper investigates the effect on corrosion resistance from the clinical washing decontamination (WD) process, specifically by studying the changes in surface roughness and Cr/Fe ratio. Results indicate that the WD process provides a positive effect on smooth polished samples, while a lesser positive effect was observed on rough reflection reduced samples.

KEYWORDS: Stainless Steel; Surgical Instrument;
Decontamination; Corrosion

INTRODUCTION:

Surgical instruments were defined as critical items by Earle H. Spaulding[1] in 1968 due to their high risk of disease transmission between patients. Both decontamination and sterilization are critical for surgical instrument reprocessing between each use. [1-3]

Corrosion on surgical instruments provides a seat for contamination, allows entrapment of debris and prevents proper sterilisation. It will place both patients and medical staff at risk of infection. Corrosion can also compromise the structural integrity of instruments and lead to mechanical failure in use. [4]

Good quality surgical instruments, with proper care, are expected to have a lifespan of more than 10 years.[4, 5] It is widely believed that incidents are under reported, yet we have still received several complaints regarding corrosion of surgical instruments, some shortly after purchase. Many factors can contribute to instruments' lifespan, including material selection, surface roughness, passivation, handling (before, during and after surgery), transportation, storage, decontamination and sterilization procedure. After purchase, the most important factors are then narrowed down to handling and the decontamination process.[4]

There is increasing concern regarding alkaline detergent used in Sterile Service Departments (SSD). According to Spry [4], strongly alkaline detergents are not recommended for routine processing as “they can destroy the passivation layer and promote corrosion”. However, alkaline detergents have been proven to be effective in

minimizing prion transmission risks [6] and this is a primary reason alkaline detergents are widely used in SSDs across the UK.

Surface treatment applied by suppliers varies greatly and results in a difference in surface roughness and surface topography. It is believed by some that a polished surface can cope with the decontamination cycle better as a mirror-like surface does not hold moisture, while others prefer reflection-reduced instruments due to their better feel and reduced light reflection.

Literature indicates that smooth surface provides better corrosion resistance by leaving less liquid residue, requiring higher pitting potential and resisting corrosion propagation than a rough surface. [7-10] However, most experiments reported were performed under aggressive environments, such as NaCl solution, and surface finishes studied were for other applications, such as food industry and construction. Therefore, these results cannot simply be assumed to apply to SSD circumstances, as there is no proof of alkaline detergent corroding instruments.

Although an oxide layer which tightly adheres to the metal and separates the metal from electrolyte solution can be provided by passivation, it remains invisible to naked eyes (usually 1-3 nanometers in thickness).[11] Since passivation cannot be tested or assured, our samples are not passivated amounting to a worst-case scenario.

In this study, effects of the WD process are discussed and the alkaline detergent is considered as part of the procedure. Roughness

(Ra) changes and Chromium/Iron (Cr/Fe) ratio were used as critical factors to determine the sample corrosion resistance.

The primary aim of this study is to investigate the effect of standard WD processes on surgical instruments and whether the passivation layer is enhanced or there is an increased risk of corrosion. The secondary aim is to investigate how the effect varies with different surface finished and hence what finish is most corrosion resistant.

METHODS:

Samples were prepared from commercial AISI 420 stainless steel in sheet form. This is one of the most widely used grades for surgical instruments, particularly for such as scissors, due to a good combination of corrosion resistance and mechanical strength.[5] Sample sheets had a thickness of 6.5mm and were cut into the dimension of 20 x 40 mm. Chemical composition is given in Table 1.

Four different surface finishes were prepared:

- a) Hand polished with 320 grit silicon carbide abrasive sheet (P320)
- b) Hand polished with 600 grit silicon carbide abrasive sheet (P600)
- c) Hand polished with 1200 grit silicon carbide abrasive sheet (P1200) and
- d) Reflection-reduced/Bead blasted with glass bead grade 10, particle size 180-300 micron (B10)

Samples were degreased in an acetone bath for 5 minutes, rinsed in deionised water for 5 minutes and dried in air.

Experiments were carried out in SSD of Ninewells Hospital, Dundee. 150 cycles were processed, mimicking one year of use. The applied decontamination procedure is described below.

Samples were aligned in a stainless steel washing basket (manufacture and material unknown due to long service period) prior to washing cycles. Each cycle includes

- 1) Pre-washed in cold water of a maximum 35°C for 6 min
- 2) Washed in 70°C water for 15 min, with detergent added
- 3) Rinsed in 65°C water for 3 min
- 4) Rinsed in 65°C water for 3 min
- 5) Final rinsed in 90°C RO water for 2min and
- 6) Hot air dried for 20 min at 120°C.

The detergent used in WD cycles is Maximum pH Plus (Serchem: Telford, UK), with a pH value of 13-14 and diluted to about pH 10.5 at point of use.

A Jeol JSM-4700F scanning electron microscope (SEM) (Jeol: Tokyo, Japan) was used at 10kV to determine the characteristic appearance of the surface finishes at 100X, 300X, 500X and 1000X.

Roughness of the samples was measured by a Dektak 3ST profilometer (Veeco: Plainview, USA) at a random direction across the surface, and a mean of 10 readings taken. Surface roughness measurements were carried out both before and after corrosion experiments.

To analyse the relationship between the concentration of chemical composition and depth from surface, depth profile analyses were

performed on samples before and after corrosion experiments by a glow discharge optical emission spectroscopy (GD-OES) (Horiba Jobin Yvon: Kyoto, Japan). The power applied was 50W and the pressure 850Pa. The sampling diameter was 15mm for each measurement.

RESULTS:

SEM images taken of different polished surface finishes (Fig.1) indicated similar morphological features among all three polished samples except fewer and shallower scratches with the finer polishing grade. The bead blasted sample surface was not characterised by scratches but indentations left after blasting.

Roughness of the sample surfaces was determined by Ra values before and after processing through the WD cycles (Table 2).

Although mean values of polished samples decreased slightly, both maximum and minimum values decreased significantly (Fig.2).

Comparing measurements before and after WD cycles, both P600 and P1200 displayed a smaller standard deviation, indicating a more uniform surface. Compared to polished samples, B10 has a minimal change in roughness as the value of Ra decreased for only 0.6% and its standard deviation covers a similar range.

It appears that the WD procedure used can smooth surgical instrument surface over time. Comparing three polished samples, it is noted that the smoother the original surface, the greater is the effect. Although the mean roughness of the sample B10 decreased the most among all four groups, it is based on a much rougher original surface.

After 150 WD cycles, all samples retained the same appearance except B10. Corrosion was observed on bead blasted samples. (Fig.3) However, it remains curious that only one side, that touching the washing basket during experiment, was corroded. It is suspected to be caused by the microenvironment formed between two materials (stainless steel sample and washing basket) because the corrosion occurred is galvanic corrosion instead of pitting corrosion.

Chromium enrichment at the surface is a critical factor for corrosion resistance. The passivation layer is mainly composed of Fe_2O_3 and Cr_2O_3 and Cr oxides are more stable and dense than Fe oxides. Hence, high chromium concentration at the material surface provides better corrosion resistance.[12] Cr/Fe ratio is often used to indicate corrosion resistance.[13, 14]

The time taken to sputter gives a measure of distance from the surface. The concentration of elements was detected before and after experiments and the Cr/Fe ratio was calculated to indicate the passivity of sample surface (Fig.4). The non-corroded area of sample B10 (Fig.4 (D)) was used in this experiment.

Among all samples, P1200 showed the most significant effect in Cr/Fe ratio increasing from 0.24 to 0.32, followed by P600, whose Cr/Fe ratio changed from 0.18 to 0.23. P320 displayed the least increase of the three polished samples with just more than 0.01. B10 has a minimal change in Cr/Fe ratio. Changes of Cr/Fe ratio caused by WD process only have an effect within a few microns of the surface (displayed on figures from 0.8s to about 1.2s), before the

Cr/Fe ratio of the material reaches back to its bulk compositional ratio.

All samples showed an increase of Cr/Fe ratio at the sample surfaces, indicating that WD cycle using alkaline detergent have a positive effect on corrosion resistance of polished stainless steel type 420. The effect ranks as P1200 > P600 > P320 > B10.

CONCLUSIONS:

By combining roughness results and depth profile of the samples, it may be concluded:

1. Currently used washing decontamination process with alkaline detergent has a positive effect on stainless steel type 420;
2. It would be helpful in extending surgical instrument life time to reduce roughness and increase the Chromium concentration at the surface;
3. The effect increases as roughness is reduced.

Inter-material reaction with stainless steel basket tray will be studied in the future, as one factor of surgical instrument corrosion problem appears galvanic corrosion.

More grades of stainless steel used to manufacturer surgical instruments, such as AISI 410 (forceps and retractors), AISI 303 (chisels and bone curettes) and AISI 316L (medical implants) will be included in future experiments. Work will also be carried in collaboration with several manufacturers using commercial surface finish grades including polishing, sand and bead blasting'.

ACKNOWLEDGEMENT:

This work was funded through support from the University of Dundee, Northern Research Partnership (NRP) and Scottish Funding Council (SFC).

REFERENCES:

1. Rutala, W.A. and D.J. Weber, *Disinfection and sterilization: an overview*. Am J Infect Control, 2013. 41(5 Suppl): p. S2-5.
2. Holmes, S., *An overview of current decontamination practices of surgical instruments and medical devices*. 2014: p. 503-547.
3. Southworth, P.M., *Infections and exposures: reported incidents associated with unsuccessful decontamination of reusable surgical instruments*. J Hosp Infect, 2014. 88(3): p. 127-31.
4. Spry, C.C., *Care and Handling of Basic Surgical Instruments*. AORN Journal, 2007. 86: p. S77-S81.
5. Mainier, F.B., et al., *corrosion in surgical instruments*. IOSR Journal of Engineering, 2013. 3(10): p. 25-31.
6. Commission, v.-O. and A. Workgroup, *Investigation of Alkaline Detergents for Automated Reprocessing of Ophthalmic Surgery Instruments in Terms of Material Compatibility and Alkaline Residues*. Zentral Sterilisation, 2009. 17(4): p. 245-256.
7. Lee, S.-J. and J.-J. Lai, *The effects of electropolishing (EP) process parameters on corrosion resistance of 316L*

- stainless steel*. Journal of Materials Processing Technology, 2003. 140(1-3): p. 206-210.
8. Hilbert, L.R., et al., *Influence of surface roughness of stainless steel on microbial adhesion and corrosion resistance*. International Biodeterioration & Biodegradation, 2003. 52(3): p. 175-185.
 9. Zatkalikova, V. and T. Liptakova, *Pitting Corrosion of Stainless Steel at the Various Surface Treatment*. Materials Engineering, 2011. 8: p. 115-120.
 10. Lee, S.M., et al., *Surface roughness and the corrosion resistance of 21Cr ferritic stainless steel*. Corrosion Science, 2012. 63: p. 404-409.
 11. Landolt, D., *Corrosion and Surface Chemistry of Metals*. first edition ed. 2007: EPFL Press. 400.
 12. Cunat, P.-J. *selection and use of stainless steels in water system*. [cited 2015 22.05.2015].
 13. Trigwell, S. and G. Selvaduray. *Effect of Surface Treatment on the Surface Characteristics of AISI 316L Stainless Steel in the Materials and Processes of Medical Devices*. 2005. Boston: ASM International.
 14. Freire, L., et al., *The electrochemical behaviour of stainless steel AISI 304 in alkaline solutions with different pH in the presence of chlorides*. Electrochimica Acta, 2011. 56(14): p. 5280-5289.

TABLES:

Table 1 Chemical composition (wt%) of sample material (AISI420)

Element	C	Si	Mn	S	P	Ni	Cr	Mo	Fe
	0.14-0.20	≤1	≤1	≤0.03	≤0.04	≤1	11.50-13.50	≤0.3	Bal.

Table 2 Roughness value of samples and changes after WD cycles

	P320	P600	P1200	B10
Before (nm)	126.0	62.5	46.3	731.4
After (nm)	122.2	59.2	43.2	727.0
Change	-3.02%	-5.27%	-6.62%	-0.60%

FIGURE CAPTIONS:

Fig.1 SEM pictures of different surface finishes (A) P320

Fig.1 SEM pictures of different surface finishes (B) P600

Fig.1 SEM pictures of different surface finishes (C) P1200

Fig.1 SEM pictures of different surface finishes (D) B10

Fig.2 Roughness value of different surface finishes (A) P320

Fig.2 Roughness value of different surface finishes (B) P600

Fig.2 Roughness value of different surface finishes (C) P1200

Fig.2 Roughness value of different surface finishes (D) B10

Fig.3 Corrosion observed on bead blasted sample after WD cycles (A)
overview

Fig.3 Corrosion observed on bead blasted sample after WD cycles (B)
x100

Fig.4 Cr/Fe value by sputtering time (A) P320

Fig.4 Cr/Fe value by sputtering time (B) P600

Fig.4 Cr/Fe value by sputtering time (C) P1200

Fig.4 Cr/Fe value by sputtering time (D) B10

FIGURES: