University of Dundee

Extra-oral appliances in orthodontic treatment

Almuzian, Mohammed; Alharbi, Fahad; McIntyre, Grant

Published in:
Dental Update

Publication date:
2016

Document Version
Peer reviewed version

[Link to publication in Discovery Research Portal](#)

Citation for published version (APA):

Almuzian, M., Alharbi, F., & McIntyre, G. (2016). Extra-oral appliances in orthodontic treatment. *Dental Update*, 43(1), 74-82.

General rights

Copyright and moral rights for the publications made accessible in Discovery Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from Discovery Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain.
- You may freely distribute the URL identifying the publication in the public portal.

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Orthodontics

Extra-oral appliances in orthodontic treatment

Mohammad Almuzian

Honorary StR in Orthodontics, Glasgow Dental Hospital & School

BDS, MFDS RCS(Edin), MFD RCS(Irel), MFDS RCPS(Glasg), MSc.HA (USA), MJDF RCS(Eng), MSc.

Fahad Alharbi

Orthodontic PhD student, Dental School, University of Dundee

Grant McIntyre

Consultant / Honorary senior lecturer in Orthodontics

BDS, FDS RCPS(Glasg), MOrth RCS(Edin), PhD, FDS(ORTH) RCPS(Glasg), FDS RCS (EDIN), FHEA

Correspondence to:

Dr G. T. McIntyre, Consultant Orthodontist, Dundee Dental Hospital and School, 2 Park Place, Dundee, DD1 4HR, UK. E-mail: grant.mcintyre@nhs.net

Extra-oral appliances in orthodontic treatment

Abstract

Extra-oral oral appliances are used in orthodontics to apply forces to the jaws, dentition or both and the popularity of these appliances is cyclical. Although the use of retraction headgear for the management of Class II malocclusion has declined over the last 20 years with the refinement of non-compliance approaches including temporary anchorage devices, headgear still has a useful role in orthodontics. The use of protraction headgear has increased as more evidence of its effectiveness for the treatment of Class III malocclusion has become available. This paper describes the mechanics and contemporary uses of headgear in orthodontics for primary care dentists and specialist orthodontists.

Clinical relevance

Extra-oral appliances have specific uses in orthodontic biomechanics. Clinicians using retraction headgear and protraction headgear should be familiar with their clinical indications, the potential problems and how these can be avoided.

Objective

To develop the knowledge of the primary care dentist and specialist orthodontist in relation to extra-oral appliances in orthodontics.

Extra-oral appliances in orthodontic treatment

Introduction

Extra-oral forces used in orthodontic treatment require the use of headgear. After the introduction of retraction headgear by Norman W. Kingsley in the late 1800s, headgear use increased with the popularisation of the first generation of fixed appliances and retraction headgear by Edward H. Angle. Retraction headgear use reduced dramatically in the 1920s with the introduction of intraoral elastics as these were believed to provide equivalent forces (1). In the 1940s, the increasing use of cephalometric radiography in orthodontics led to concerns about the adverse effects of intraoral elastic traction including proclination of lower incisors and retroclination of upper incisors. As a result, headgear use increased again. Although the use of retraction headgear has declined steadily with the introduction of non-compliance fixed appliance distalising appliances including the pendulum appliance, lip bumper appliances and temporary anchorage devices (TADs), headgear remains the standard for anchorage reinforcement and maxillary arch distalisation. Enthusiasm is increasing for the use of protraction headgear for the treatment of Class III malocclusion and patients affected by hypodontia where space closure may result in a reverse overjet.

There are a variety of clinical uses for headgear, requiring different directions of forces, force levels and wear levels which result in varying effects on the dentition and/or skeletal base relationships. In this article, the contemporary uses and effectiveness of extra-oral traction appliances are discussed.

There are three main components:

1. **Extraoral unit:** This part of the appliance provides the anchorage for the extra-oral force in a form of headcap, neckpad/strap or chin cup (for retraction headgear) (Figure 1) or a facemask (for protraction headgear). Retraction headgear normally includes the delivery force system in the headcap / neckpad/strap.

2. **Force delivery system:** this can involve either a spring loaded device (Figure 2) or heavy-force extra-oral elastic. The former is used with retraction headgear and chin cup appliances while the latter is used with protraction facemask appliances.
3. **Intermediate / connecting component:** This transmits the force to the teeth and underlying skeleton and connects the extra-oral and intraoral components. With retraction headgear, the intermediate component is the outer part of the facebow that connects to the extra-oral component. The inner bow should also incorporate a robust safety mechanism to prevent a accidental disengagement (2, 3) (Figure 3a,b).
4. **Intraoral component:** Retraction headgear is attached via the headgear facebow to either fixed, removable or functional appliances for either anchorage reinforcement or molar distal movement. With a fixed appliance (Figure 4), the inner bow inserts to molar bands using welded headgear tubes which are either gingivally placed (close to the centre of tooth rotation) or occlusally positioned (for better access during insertion and removal). With removable appliances (Figure 5), the facebow either inserts into tubes soldered to the molar clasps or to coils wound as part of the clasps (Figure 3b). Headgear tubes can also be incorporated into the acrylic of a functional appliance for mid-arch orthopaedic force delivery. The only headgear appliance that has no intraoral component is the chin-cup appliances (4). When using a protraction facemask, the intraoral component is either a fixed or removable appliance with many operators simultaneously treating the patient with slow maxillary expansion (using a quad helix appliance) or rapid maxillary expansion (RME) (5, 6).

Uses of extra-oral appliance in clinical orthodontics

The effectiveness of headgear depends on entirely patient compliance in wearing the appliance and it is well-known that patients usually overestimate the wear of headgear. (7) As headgear is not worn full-time, the intermittent force delivery is a biomechanical disadvantage with relapse occurring when the appliance is not worn. Headgear charts, timers and award systems have been proposed to improve compliance in order to improve

success rates. (8) There are different level of force and duration of use according to the clinical application. (Table 1)

1. Anchorage reinforcement

Retraction headgear can be used for anchorage reinforcement antero-posteriorly and vertically. Although headgear is considered to be a good method of anchorage reinforcement, TADs have been shown to provide equivalent results with minimal patient co-operation (9). For anchorage reinforcement, force levels of 250-300g per side are required for a minimum of 10 hours/day, on average. (10-14).

2. Molar Distalisation:

Retraction headgear can distalise the molars by up to ½ molar unit (approximately 5mm) which can be useful used for correction of a ½ unit Class II molar relationship and relief of crowding on a non-extraction basis or reduction of an increased overjet. (15-18). Molar distalisation using headgear in conjunction with extraction of the deciduous canine in order to provide space for ectopic permanent maxillary canines has been shown to have a success rate of 80% in comparison to 50% where the deciduous canine was extracted as an isolated measure (19). However, the Cochrane review by Parkin et al has shown there is no high-quality evidence to justify the interceptive effect of extraction of primary canines (20). Headgear can also be used to upright impacted upper first molars and to regain lost space after premature loss of primary teeth. (21). Force levels of 400-500g/side used over a period of 12-14 hours/day are necessary for distalisation. (10) Headgear use can tip and extrude the molars depending on the magnitude and direction of the applied force as well as the duration of headgear wear. (11-14).

3. *Canine retraction/labial segment movement:*

J-hook headgear has historically been used to intrude upper anterior teeth, retract maxillary canines individually or rarely, retract lower canines. (22). Force levels of 250-300g per side are required for a minimum of 10 hours/day, on average. (10-14) Due to the risk of ocular trauma and as canine retraction is more easily achieved with sliding mechanics when using contemporary fixed appliances, J-hook headgear is no longer used.

4. *Differential (asymmetric) tooth movement:*

Asymmetric movement of the molars can be achieved using asymmetrical headgear (AHG) (23). There are many designs for AHG but the main principle of action is Castagliano's Theorum which involves greater movement on one the side due to a longer outer bow or heavier force on this side. The variety of designs of AHG include the power-arm facebow, soldered-offset face-bow, swivel-offset face-bow and spring-attachment face-bow (24-26). One of the disadvantages of AHG is the tendency to produce a scissor bite on the side of the long arm and an increase in difficulty when fitting the appliance (27). Force levels of 250-300g per side are required for a minimum of 10 hours/day, on average. (10-14)

5. *Growth modification*

- a. Class II malocclusion: theoretically, headgear can alter the skeletal relationship, due to a restriction of maxillary forward and downward growth that allows the mandible to "catch up" during treatment, when used in growing children aged 12-13 years (28, 29). However, most of the skeletal improvement has been shown to be lost 1 year after treatment. (28, 30)
- b. Class III malocclusion: Enthusiasm for the use of protraction facemask headgear for the treatment of cases with a Class III incisor relationship is increasing. Studies have found that the facemask may provide an orthopaedic effect in growing patients with an improvement in the ANB angle by an average of 2.6 degrees when compared to a control group. (31) Unlike the loss of the skeletal improvement with retraction headgear in Class II cases over time, the

skeletal results with protraction headgear are generally stable after 3 years of treatment. (31, 32) Although, investigators have found that relapse can occur up to 8.5 years after treatment (33) protraction facemask is now the treatment of choice for mild to moderate Class III malocclusions (34).

c. High-pull headgear combined with a twin block, monoblock or Dynamax appliance can be used to treat high angle class II malocclusions. (35) However, the evidence for the effect of headgear on the vertical dimension is weak as Oliveira et al found only a limited improvement in anterior open bite when treated with a palatal crib and headgear. (36). Force levels of 400-500g/side used over a period of 12-14 hours/day are necessary for a skeletal effect (10).

Headgear types

Headgear is categorized according to the antero-posterior direction of pull: if the force is directed distally, then the headgear called retraction and where the force is mesially directed, it is called protraction headgear. Retraction headgear is also classified according to the direction of force above, at, or below the occlusal plane: high pull, straight/combination pull and low or cervical pull.

I. Retraction headgear

1. *Low-pull or cervical traction (figure 6a)*: This type of headgear is frequently referred to as cervical traction and is the most commonly used headgear appliance. (37) Cervical traction is used mainly for the correction of Class II malocclusion by restraining the forward and downward growth of the maxilla. (38) Cervical traction is believed to have an interaction with the growth of the mandible as well as extrusion of maxillary molars. (39) This latter effect results in a clockwise mandibular rotation thus cervical headgear is indicated mainly for growing children with a deep overbite and is contra-indicated in cases with vertical skeletal discrepancies and anterior open bites. (39, 40)

2. *Straight/combination pull (figure 6b)*: This is a hybrid of high pull and low pull, producing a pure distal movement without any extrusion or intrusion of the molars. (37) The forces are transmitted to the teeth through a combination of a head cap and a neck strap. (23)
3. *High occipital pull (figure 6c)*: High pull headgear produces forces that pass apically through the centre of resistance of the maxillary teeth and clinical investigations have demonstrated that high pull headgear can perform distal movement of the molars effectively. (4) With high pull headgear, it is possible to produce intrusive forces to the molars rather than extrusive forces, which can therefore help the correction of an anterior open bite. (39) Occipital headgear can also produce orthopaedic effects on the maxilla by restraining the vertical growth of the maxilla. (37) Although wearing high pull headgear can result in compensatory eruption of the mandibular molars, this is can be controlled by using a fixed lingual arch.

II. J-Hook:

1. *High pull*: This type of headgear exerts an intrusive and distal action on the upper incisors which moves the teeth distally as well as gingivally. (37)
2. *Straight/combination pull*: This is used to retract mandibular or maxillary canines distally. However, the force vector can cause downward tipping of the incisors. (37)
3. *Low Pull*: This type of headgear extrudes and retracts the mandibular incisors to camouflage a Class III skeletal base relationship. (37). It is worth noting that the use of J-Hook headgear in the UK is obsolete due to safety concerns about potential ocular injuries.

III. Protraction Headgear (figures 7,8)

This is also called a facemask or reverse headgear. Facemask therapy is an effective orthopaedic appliance for growing children in order to correct a Class III malocclusion through forward movement of the maxilla. (41) The results of treatment are better in the early mixed dentition than in the late mixed dentition. (42) In addition to the skeletal changes,

reverse headgear can result in dental compensation to assist with the correction of a reverse overjet or Class III malocclusion. (14, 43)

IV. Chin-Cup

Chin-cup therapy is used to correct Class III malocclusions in growing patients using forces ranging from 250-600 g per side for at least one year. Many researchers have found that chin-cup therapy has effects on mandibular protrusion through dental effects as well as redirecting, inhibiting or slowing condylar growth. (44, 45) However, as no forces are applied to the maxilla, cases with maxillary hypoplasia will not have any upregulation of maxillary growth when treated with a chin-cup appliance. As a result, chin-cup therapy is seldom used in the UK.

Fitting and monitoring progress with headgear

Retraction

The first step in fitting the retraction headgear involves selecting the correct facebow size. The inner bow should be 1.13mm while the outer bow is 1.45mm for maximum rigidity. The facebow should be set parallel to occlusal plane with slight expansion. The centre of bow should be slightly away and above the central incisor edge. The inner bow should follow the contour of lip and cheeks but not actively displacing them. Each side of the facebow must be adjusted to fit into the fixed appliance molar bands or removable appliance tubes/colis at one side at a time. If the molar are rotated then toe-in bends should be placed. The length of the outer bow, relationship to the centre of rotation, the direction of pull should be carefully adjusted to minimise distal crown tipping. All these factors are important in determining the force vectors and the net force. In particular, if the outer bow is positioned at the trifurcation point of molars, the result will be distal translocation of the molars (10,46).

Protraction

With protraction headgear, the fitting procedure starts with adjusting the cams (Figures 7,8) using the Allen key until they are 15 degrees below the occlusal plane. The elastics are

attached to the intraoral device and should be heavy enough to apply the required force at 30 degree angulation. It is preferable for the elastics to cross-over to avoid lip irritation. (31)

Monitoring progress with headgear appliances can be undertaken by asking the patient / parent about compliance, using compliance charts (8), assessing the ability of the patient to insert/remove the appliance, checking for physical signs of wear and tear, identifying positive tooth movement in comparison with pre-treatment study models/ cephalometry and detecting molar mobility.

Iatrogenic effects of retraction and protraction headgear and solutions to overcome these

There are many iatrogenic effects of extra-oral appliances. These are detailed in Table 2 with solutions that can be used to manage them.

Safety mechanisms for retraction headgear

Many safety mechanisms are available to prevent ocular injuries resulting from retraction headgear. These include:

1. Safety release mechanisms where the headgear is designed to 'break-away' when excessive force is applied (Figure 2). (56)
2. Safety facebows such as locking mechanisms (Figure 3a) (Nitom Locking Facebow, Ortho Kinetics Corporation, Vista, Calif/GAC International Inc, Central Islip, NY) and recurved reverse entry inner bows. (2, 3)
3. Additional safety mechanics such as blunt ends and locating elastics. (2)
4. Masei (www.masei.com) safety strap (rigid neck strap). (2)

The British Orthodontic Society (www.bos.org.uk) recommends at least two main safety mechanics are provided with each headgear appliance as well as informed, written and

verbal instructions being given to the patient. Along with these instructions, the patient should be given a clear demonstration of how to insert and remove the headgear. It is vital to inform the patient to ensure the safety mechanisms are in place during use. Patients should be advised to avoid wearing their headgear while playing sports and they should stop the use of the headgear and contact their orthodontist immediately if the headgear becomes detached during sleep. Moreover, the patient must be aware that if any eye injury associated with the headgear occurs, it must be treated as a medical emergency. Lastly, patients should be instructed to bring their headgear to each appointment and report any problems to their orthodontist. (51)

Conclusion

Retraction headgear remains a useful appliance in contemporary clinical orthodontics. Protraction headgear has both an orthopaedic and orthodontic effect and is a useful appliance for young patients presenting with a Class III malocclusion.

Acknowledgement

Mr Simon Scott, Medical Photography, University of Dundee is thanked for his help with the production of the images.

References

1. Angle EH. Treatment of malocclusion of the teeth: Angle's system: Philadelphia: White Dental Manufacturing Company, 1907.
2. Postlethwaite K. The range and effectiveness of safety headgear products. *Eur J Orthod* 1989; 11: 228-234.
3. Samuels R, O'Neill J, Bhavra G, Hills D, Thomas P, Hug H, et al. A clinical evaluation of a locking orthodontic facebow. *Am J Orthod Dentofacial Orthop* 2000; 117: 344-350.
4. Graber LW. Chin cup therapy for mandibular prognathism. *Am J Orthod* 1977; 72: 23-41.
5. Baek S-H, Kim K-W, Choi J-Y. New treatment modality for maxillary hypoplasia in cleft patients. *Angle Orthod* 2010; 80: 783-791.
6. Delaire J. Maxillary development revisited: Relevance to the orthopaedic treatment of Class III malocclusions. *Eur J Orthod* 1997; 19: 289-311.
7. Brandão M, Pinho HS, Urias D. Clinical and quantitative assessment of headgear compliance: A pilot study. *Am J Orthod Dentofacial Orthop* 2006; 129: 239-244.
8. Cureton SL, Regennitter FJ, Yancey JM. The role of the headgear calendar in headgear compliance. *Am J Orthod Dentofacial Orthop* 1993; 104: 387-394.
9. Kuroda S, Yamada K, Deguchi T, Kyung H-M, Takano-Yamamoto T. Class II malocclusion treated with miniscrew anchorage: comparison with traditional orthodontic mechanics outcomes. *Am J Orthod Dentofacial Orthop* 2009; 135: 302-309.
10. Bowden DE. Theoretical considerations of headgear therapy: a literature review. 2. Clinical response and usage. *Br J Orthod* 1978; 5: 173-181.
11. Taner TU, Yukay F, Pehlivanoglu M, Çakırer B. A Comparative Analysis of Maxillary Tooth Movement Produced by Cervical Headgear and Pend-X Appliance. *Angle Orthod* 2003; 73: 686-691.
12. Tanne K, Matsubara S, Sakuda M. Stress distributions in the maxillary complex from orthopedic headgear forces. *Angle Orthod* 1993; 63: 111-118.

13. Tortop T, Yüksel S. Treatment and Posttreatment Changes with Combined Headgear Therapy. *Angle Orthod* 2007; 77: 857-863.
14. Wells AP, Sarver DM, Proffit WR. Long-term Efficacy of Reverse Pull Headgear Therapy. *Angle Orthod* 2006; 76: 915-922.
15. Atherton G, Glenny A-M, O'Brien K. Development and use of a taxonomy to carry out a systematic review of the literature on methods described to effect distal movement of maxillary molars. *J Orthod* 2002; 29: 211-216.
16. Firouz M, Zernik J, Nanda R. Dental and orthopedic effects of high-pull headgear in treatment of Class II, division 1 malocclusion. *Am J Orthod Dentofacial Orthop* 1992; 102: 197-205.
17. Melsen B, Dalstra M. Distal molar movement with Kloeohn headgear: Is it stable? *Am J Orthod Dentofacial Orthop* 2003; 123: 374-378.
18. Proffit WR, Fields Jr HW, Sarver DM. Contemporary orthodontics (4th ed). St Louis: Elsevier Health Sciences, 2006.
19. Leonardi M, Armi P, Franchi L, Baccetti T. Two interceptive approaches to palatally displaced canines: a prospective longitudinal study. *Angle Orthod* 2004; 74: 581-586.
20. Parkin N, Furness S, Shah A, Thind B, Marshman Z, Glenroy G, et al. Extraction of primary (baby) teeth for unerupted palatally displaced permanent canine teeth in children. *Cochrane Database Syst Rev* 2012; 12: CD004621.
21. Kurol J, Bjerklin K. Treatment of children with ectopic eruption of the maxillary first permanent molar by cervical traction. *Am J Orthod* 1984; 86: 483-492.
22. Perez CA, de Alba JA, Caputo AA, Chaconas SJ. Canine retraction with J hook headgear. *Am J Orthod* 1980; 78: 538-547.
23. Holmes A, Nashed R, O'Keeffe C. The correction of dental centre line discrepancies using an edgewise appliance. *J Orthod* 1989; 16: 271-276.
24. Hershey HG, Houghton C, Burstone CJ. Unilateral face-bows: a theoretical and laboratory analysis. *Am J Orthod* 1981; 79: 229-249.

25. Brosh T, Portal S, Sarne O, Vardimon AD. Unequal outer and inner bow configurations: Comparing 2 asymmetric headgear systems. *Am J Orthod Dentofacial Orthop* 2005; 128: 68-75.
26. Jacobson A. A key to the understanding of extraoral forces. *Am J Orthod* 1979; 75: 361-386.
27. Martina R, Viglione G, Teti R. Experimental force determination in asymmetric facebows. *Eur J Orthod* 1988; 10: 72-75.
28. Keeling SD, Wheeler TT, King GJ, Garvan CW, Cohen DA, Cabassa S, et al. Anteroposterior skeletal and dental changes after early Class II treatment with bionators and headgear. *Am J Orthod Dentofacial Orthop* 1998; 113: 40-50.
29. Tulloch JFC, Phillips C, Koch G, Proffit WR. The effect of early intervention on skeletal pattern in Class II malocclusion: A randomized clinical trial. *Am J Orthod Dentofacial Orthop* 1997; 111: 391-400.
30. Tulloch J, Phillips C, Proffit WR. Benefit of early Class II treatment: progress report of a two-phase randomized clinical trial. *Am J Orthod Dentofacial Orthop* 1998; 113: 62-74.
31. Mandall N, DiBiase A, Littlewood S, Nute S, Stivaros N, McDowall R, et al. Is early Class III protraction facemask treatment effective? A multicentre, randomized, controlled trial: 15-month follow-up. *J Orthod* 2010; 37: 149-161.
32. Mandall N, Cousley R, DiBiase A, Dyer F, Littlewood S, Mattick R, et al. Is early Class III protraction facemask treatment effective? A multicentre, randomized, controlled trial: 3-year follow-up. *J Orthod* 2012; 39: 176-185.
33. Masucci C, Franchi L, Defraia E, Mucedero M, Cozza P, Baccetti T. Stability of rapid maxillary expansion and facemask therapy: A long-term controlled study. *Am J Orthod Dentofacial Orthop* 2011; 140: 493-500.
34. Watkinson S, Harrison J, Furness S, Worthington H. Treatment for prominent lower front teeth in children. *Cochrane Database Syst Rev* 2013; 9: CD003451.

35. Parkin NA, McKeown HF, Sandler PJ. Comparison of 2 modifications of the Twin-block appliance in matched Class II samples. *Am J Orthod Dentofacial Orthop* 2001; 119: 572-577.
36. Lentini-Oliveira D, Carvalho F, Qingsong YE, Junjie L, Saconato H, Machado M, et al. Orthodontic and orthopaedic treatment for anterior open bite in children. *Cochrane Database Syst Rev* 2007; 2: CD005515.
37. Bowden DE. Theoretical considerations of headgear therapy: a literature review. 2. Clinical response and usage. *Br J Orthod* 1978; 5: 173-181.
38. Poulton DR. The influence of extraoral traction. *Am J Orthod* 1967; 53: 8-18.
39. Barton JJ. High-pull headgear versus cervical traction: A cephalometric comparison. *Am J Orthod* 1972; 62: 517-529.
40. Wieslander L, Tandlåkare L. The effect of orthodontic treatment on the concurrent development of the craniofacial complex. *Am J Orthod* 1963; 49: 15-27.
41. Turner P. Extra-oral traction. *Dent Update* 1991; 18: 197-203.
42. Baccetti T, McGill JS, Franchi L, McNamara Jr JA, Tollaro I. Skeletal effects of early treatment of Class III malocclusion with maxillary expansion and face-mask therapy. *Am J Orthod Dentofacial Orthop* 1998; 113: 333-343.
43. Ngan P, Hägg U, Yiu C, Merwin D, Wei SH. Soft tissue and dentoskeletal profile changes associated with maxillary expansion and protraction headgear treatment. *Am J Orthod Dentofacial Orthop* 1996; 109: 38-49.
44. Deguchi T, Kuroda T, Minoshima Y, Graber TM. Craniofacial features of patients with Class III abnormalities: Growth-related changes and effects of short-term and long-term chin cup therapy. *Am J Orthod Dentofacial Orthop* 2002; 121: 84-92.
45. Abdelnaby YL, Nassar EA. Chin cup effects using two different force magnitudes in the management of Class III malocclusions. *Angle Orthod* 2010; 80: 957-962.
46. Yoshida N, Jost-Brinkmann P-G, Yamada Y. Initial tooth movement under extraoral force and considerations for controlled molar movement. *Angle Orthod* 1995; 65: 199-208.

47. Deguchi T, Murakami T, Kuroda S, Yabuuchi T, Kamioka H, Takano-Yamamoto T. Comparison of the intrusion effects on the maxillary incisors between implant anchorage and J-hook headgear. *Am J Orthod Dentofacial Orthop* 2008; 133: 654-660.
48. Samuels RH, Willner F, Knox J, Jones ML. A national survey of orthodontic facebow injuries in the UK and Eire. *Brit J Orthod* 1996; 23: 11-20.
49. Holland GN, Wallace DA, Mondino BJ, Cole SH, Ryan SJ. Severe ocular injuries from headgear. *Am J Orthod* 1986; 89: 173.
50. Samuels RH, Willner F, Knox J, Jones ML. A national survey of orthodontic facebow injuries in the UK and Eire. *J Orthod* 1996; 23: 11-20.
51. Use of Headgear and facebows advice sheet: British Orthodontics Society 2013 [cited 2013 25/09/2013]. Available from: www.bos.org.uk.
52. Kerosuo HM, Dahl JE. Adverse patient reactions during orthodontic treatment with fixed appliances. *Am J Orthod Dentofacial Orthop* 2007; 132: 789-795.
53. Travess H, Roberts-Harry D, Sandy J. Orthodontics. Part 6: Risks in orthodontic treatment. *Brit J Orthod* 2004; 196: 71-77.
54. Rahilly G, Price N. Current products and practice nickel allergy and orthodontics. *J Orthod* 2003; 30: 171-174.
55. Nickel allergy in orthodontics 2013 [cited 2013 25/09/2013]. Available from: www.bos.org.uk.
56. Stafford GD, Caputo AA, Turley PK. Characteristics of headgear release mechanisms: Safety implications. *Angle Orthod* 1998; 68: 319-326.

Table 1: Force Duration and levels with extra-oral appliances

	Force levels (gram/side)	Duration (hour/per day)
Anchorage	250-300	10
Tooth movement	300-350	12-14
Orthopaedic movement	450-600	14-16

Table 2: Iatrogenic effects, frequency and solutions

Iatrogenic effect	Frequency	Management
Pain due to heavy force levels	Common	<ul style="list-style-type: none">• Non-steroidal painkillers
Increased risk of root resorption	Rare possibly with J-hook headgear	<ul style="list-style-type: none">• Monitor radiographically using long cone periapical radiographs.(47) F• Force levels should be as low as possible over a short duration
Trauma to the face and eye	Rare but serious consequences (ophthalmitis and blindness) (48, 49) due to accidental disengagement or recoiling injuries (50)	Prevention through: <ul style="list-style-type: none">• Demonstration of the safe use of the headgear to the patient and parent• Verbal and written instructions (51)• Incorporation of safety mechanisms
Nickel allergy Contact dermatitis-type IV (delayed hypersensitivity)	30% of female and 3% of male patients undergoing orthodontic treatment. (52)	<ul style="list-style-type: none">• Confirmation of the nickel allergy by a Dermatologist• Further episodes avoided using Nickel-free appliances or a plastic shield for the headgear facebow. (53-55).
Latex allergy	Infrequent	<ul style="list-style-type: none">• Latex-free elastic components

Figure captions

1. Figure 1: Headcap / neckstrap for retraction headgear
2. Figure 2: Force-delivery module (snap-away) for retraction headgear (open and closed)
3. Figure 3: Locking facebow a) closed b) attached to clasp of upper removable appliance
4. Figure 4: Retraction headgear fitted to a fixed appliance
5. Figure 5: Retraction headgear fitted to a removable appliance
6. Figure 6 a) straight/combination pull, b) high-pull and c) low-pull retraction headgear
7. Figure 7: Facemask (Petit type)
8. Figure 8: Facemask (Delaire type) on a) dry skull with b) rapid maxillary expansion (RME)