University of Dundee

Automated Identification from Dental Data (AutoIDD)

Reesu, Gowri Vijay; Woodsend, Brenainn ; Manica, Scheila; Revie, Gavin F.; Brown, Nathan L.; Mossey, Peter A.

Published in:
Forensic Science International

DOI:
[10.1016/j.forsciint.2020.110218](https://doi.org/10.1016/j.forsciint.2020.110218)

Publication date:
2020

Licence:
CC BY-NC-ND

Document Version
Peer reviewed version

[Link to publication in Discovery Research Portal](#)

Citation for published version (APA):

Reesu, G. V., Woodsend, B., Manica, S., Revie, G. F., Brown, N. L., & Mossey, P. A. (2020). Automated Identification from Dental Data (AutoIDD): A New Development in Digital Forensics. *Forensic Science International*, 309, 1-9. [110218]. <https://doi.org/10.1016/j.forsciint.2020.110218>

General rights

Copyright and moral rights for the publications made accessible in Discovery Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from Discovery Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain.
- You may freely distribute the URL identifying the publication in the public portal.

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Title: Automated Identification from Dental Data (AutoIDD): A New Development in Digital Forensics.

ABSTRACT:

There has been a significant expansion in the use of 3-dimensional (3D) dental images in recent years. In the field of forensic odontology, an automated 3D dental identification system could enhance the identification process. This study presents a novel method for automated human dental identification using 3D digital dental data by utilising a dental identification scenario. The total study sample was divided into two groups: Group A (120 dental models) and Group B (120 Intra-oral scans-IOS). Group A data was composed of 3D scanned dental models of post-orthodontic treated patients (30 maxillary and 30 mandibular). This data was considered as AM digital data. To generate an identical sample, the dental casts (60) of the same patients were retrieved and laser scanned. These models were considered as PM digital data. Group B data (IOS) was obtained from 30 study participants. To reconstruct a dental identification scenario 30 maxillary and 30 mandibular IOS were obtained from 30 participants and were considered as IOS-AM. After one year, another set of IOS (60) were acquired from the same participants and were considered as IOS-PM. The results showed that the AutoIDD (Automated Identification from Dental Data) software was consistent in accuracy; capable of differentiating “correct matches” (high match percentage) from “non-matches” (very low percentage) by 3D image superimposition. The match percentage of the maxillary and mandibular IOS ranged from 64-100% and 81-100%, with a mean percentage of 96.7 and 96.4 respectively. This study demonstrated the feasibility of using 3D scans through a new automated software - AutoIDD in digital forensics to assist the forensic expert in confirming the identity of a deceased individual from the available AM dental records.

KEYWORDS: Forensic odontology; Automated Identification; Dental data; 3-dimensional models; Intra-oral scans.

1. Introduction

Forensic human identification by dental means has proved to be extremely useful and reliable over the years. This can be achieved with the availability of accurate and comprehensive ante-mortem (AM) dental records which are essential for a positive identification [1,2]. Several dental data coding systems have been proposed for the use in reports and computer-assisted identifications in the field of forensic odontology. The most well-known identification applications in the electronic management of dental records were CAPMI [3], WinID3 [4], Disaster And Victim Identification ‘‘DAVID’’ [5], and the Plass Data system - The DVI System International [6]. All have been recommended by INTERPOL to its member countries as they share a common purpose. These systems have primarily automated the text searching of records and require manual processing of data.

Dental radiographs are one of the key components of dental records which assists in the process of human identification. Studies on semi-automatic [7,8] and automatic dental identification based [9] on two-dimensional (2D) radiographs have been proposed over the last two decades. However, 2D radiograph based approaches had several limitations. The tooth segmentation process was time consuming and inaccurate due to the low image quality from blurred dental radiographs. Distortions in tooth shape and arch arising from different imaging angles are significant, which renders automated 2D radiographic retrieval and identification inaccurate [8,10,11].

To overcome the inherent limitations of 2D based methods, a feasible and an efficient automated 3D dental identification system that would enhance the identification process is essential. The application of 3-dimensional (3D) imaging in dentistry has widely expanded in recent years. As a result, clinical practices and laboratory techniques are shifting to digital workflows [12]. The introduction of intra-oral scanners for direct digitalisation of the patient’s dental arches were developed as an alternative to the use of conventional impression materials [13,14]. This digital system has important advantages in reducing impression time, patient burden, efficient storage and retrieval, higher accuracy, rapid access to 3D diagnostic information, and easy transferability of digital data [15,16]. Using a laser scanner, dental casts can be converted into 3D models that may be used in alignment and matching which can be subjected to automated comparative dental analysis [17].

This study proposes a new automated system in digital forensics to assist the forensic expert in accurately identifying the correct dentition of the deceased individuals using AutoIDD (Automated Identification from Dental Data), from the available AM dental records. The objective was to test the performance of the AutoIDD software using AM and PM (post-mortem) 3D scans for accuracy and validity.

2. Materials and Method

This study was approved by the East of Scotland Research Ethics Service, REC reference: 17/ES/0144. A new automated software was designed that uses a combination of techniques including Iterative Closest Point (ICP) and Principal Component Analysis (PCA) for accurate identification using 3D dental models and intra-oral scans (IOS). An overview of the automated alignment and matching process is shown in Fig. 1.

2.1 Data Acquisition

The total study sample consisted of 240 3D maxillary and mandibular dental data. The data was divided into two groups: Group A (60AM & 60PM =120 dental models) and Group B (60AM & 60PM =120 IOS).

The data collected for Group A consisted of 60 dental casts of 30 patients (30 maxillary and 30 mandibular dental models) which were all of post-orthodontic patients. According to the clinical data storage protocol, all the patients' dental casts, who had their treatment completed, were laser scanned to create indirect 3D digital images of dental models by the laboratory technicians for digital storage. The dental casts were scanned using R700 3Shape Orthodontic Study Model Scanner (Copenhagen, Denmark) and the images were saved in stereolithography (.STL) format. Patients who had consented for their records to be used for research purposes were identified through the clinic registry by the principal investigator (PI). All the patient identifying information was removed and a unique study code was assigned to each 3D dental model by the PI. For the purposes of this study these 3D models were considered as **"AM digital data"**.

To generate an identical sample, 60 dental casts (30 maxillary and 30 mandibular) of the same patients were retrieved and laser scanned by the PI and were considered as **"PM digital data"**. Any dental casts that were found to be damaged during the storage process were excluded from the AM and PM data. This is to determine whether different scans obtained from the same patient's casts were identifiable. The rationale for using the post-orthodontic data was to investigate the performance of the software in identifying the correct AM-PM 3D dental model matches within an orthodontic treated sample.

The data collected for Group B consisted of 120 IOS in total. The dental school staff and students were invited to participate in this research, where 30 participants were recruited. To reconstruct a dental identification scenario, 30 maxillary and 30 mandibular dental arches scans were obtained using 3Shape TRIOS Intraoral Scanner (Copenhagen, Denmark) by the PI and were considered as IOS-AM. After one year, another set of IOS (60) were acquired from the same participants and were considered as IOS-PM. The purpose of this process was to determine the sensitivity of AutoIDD software towards any variations in an individual's dentition; morphological and restorative features.

In total, the two groups of data comprised of four types of data sets (two sets of maxillary and mandibular 3D dental models and IOS). On examination of the AM and PM IOS data, the following dental characteristics were observed; restorations, missing tooth, occlusal cavities-unrestored and anterior crowding of teeth.

The study hypothesis is that an accurate dental identification can be facilitated through the use of AutoIDD software. The identification process is based on dental characteristics, shape and alignment of the 3D dental arches and any dental treatment interventions which can comprise a unique set.

2.2 3D dental scans - STL files

All the scans were in .STL format (standard open-source 3D file format). Once read from file a .STL is often referred to as a 'mesh'. It is composed of numerous triangles (Fig. 2) where each triangle is represented by the (x, y, z) values of each corner. Rather than considering the corners of the triangles, AutoIDD software looks at the centre of each triangle. This is mainly for the benefit of speedy process and for the convenience of having one point per outward normal.

Fig. 2. 3D Mesh of a 3D scanned dental model.

Fig. 1. An overview of the automated alignment and matching process of AutoIDD.

2.3 Pre-processing

For an efficient functioning of the Iterative Closest Point (ICP) algorithm, pre-processing of the 3D scans is required. It is a two-step process which involves the pre-alignment and segmentation/cropping of each of the maxillary and mandibular arches, executed with the help of Principle Component Analysis (PCA). The entire collection of the given models can be pre-processed and the output is saved in advance.

2.3.1 Pre-alignment

This process is designed to consistently position and orientate any dental scan so that the centre of mass of the model is at the origin.

An overview of the steps involved:

- Use PCA to obtain three basis vectors (x, y, & z).
- Check the signs of each basis vector.
- Fit a plane through all the tips of the teeth to gain a more accurate “up” basis vector.
- Adjust the “forwards” vector to maintain orthogonality with the new “up”.

2.3.2 Principle Component Analysis

The most well-known approach computing the alignment of 3D objects is the PCA method [18]. The goal was to find a method that best aligns any 3D dental scan and will consequently align two similar 3D scans in the same way. The 3D scan has symmetries and it is aligned with particular axes or symmetry planes [19]. So the output of PCA on a dental scan should yield up/down (Z-axis) as the first unit vector, forwards/backwards (Y-axis) as the second and left/right as the third (X-axis). The occlusal surfaces were compared to the PCA’s “up” according to the dental scan, i.e. maxillary or mandible which is determined by the filename (AM/PM-Maxi.stl, AM/PM-Mand.stl) and adjusted if needed. The comparison of unit vectors is completed using the scalar product. The AutoIDD software uses the following basis vectors after a series of computations:

- The “up” (regardless of it being maxillary or mandibular arch) vector of the scan is parallel to the z-axis.
- The “right” (patient’s right) vector of the model is parallel to the x-axis.
- “Forwards” is parallel to the y-axis.

2.3.3 Segmentation

Segmentation, also referred to as cropping is defined as the process of separating the 3D model into segments that are representative of the model shape, calculated by edge detection. The goal of segmentation is to simplify the representation of a model/mesh into something that is more tangible and easier to analyse [20]. Manual segmentation is very time-consuming for large datasets to eliminate the non-dental/ plaster part of the dental model which does not contain information of the teeth. Therefore, many automated [21] and semi-automated [22] computer based systems have been developed that can accurately demarcate the desired dental component in the model.

In this software, once the pre-alignment of the model was attained, an automated segmentation plane dissecting the dental cast was executed. In Python™, which is an open source programming language (Python Software Foundation, version 3.8.0, Beaverton, USA) [23], arch meshes can be sampled according to a specific instruction i.e. “to include all points above a certain height”. AutoIDD finds the highest point in the direction [0, 1, 1] if mandibular mesh or [0, 1, -1] if maxillary mesh. This point should approximately be the midpoint of the incisal edge. Then the mesh was cropped so that anything more than 9 mm below or 45 mm behind the incisal edge midpoint was eliminated (Fig. 3). Additionally the test model was cropped by another 1mm vertically and 3mm horizontally just before the ICP process. This step was to ensure that the test scan includes areas that the reference scan contain and make certain that each point on the test scan is paired with the corresponding point on the reference model.

Standardisation of the methodology is an important element of minimising soft tissue influence and increasing accuracy. Hence, the height and length of the segmentation plane was chosen with an intension to include just the dental components in the scan. These values can be adjusted to obtain a desired plane with a goal to limit gum area. The same segmentation procedure was applied to the intra-oral scans.

Fig. { SEQ Figure * ARABIC }. Images of the 3D dental models before (a, c) and after (b, d) after the

2.4 Iterative Closest Point (ICP)

A method for an accurate and computationally efficient registration of 3D shapes was developed based on the ICP algorithm which is to find the closest point on a geometric entity to a given point [24]. The algorithm finds 3D correspondences between two point sets (a reference and a test set) and tries to determine the best match on the “reference” set in terms of minimum distance. This whole process is considered as one iteration and it continues until the alignment of the models is deemed either satisfactory or unlikely to improve.

Generally, a well matching model pair will converge quickly in 10 - 20 iterations, then improvement rates will rapidly drop. Whereas a poorly matched pair will gradually decline down to a minimum error. To give a good balance between performance and speed, AutoIDD will continue to iterate until the errors cease to improve by

at least 0.1% or when the iterations reach 100. ICP moves the test model in small steps per iteration and will always seek to make an improvement on every iteration. It is of utmost importance that the plaster base of the dental models and other non-dental features are removed and the meshes are trimmed to consistent measurements. Failure to do so leads to misleadingly high error scores and instability within ICP.

The ICP aims to find the ‘best’ superimposition of a ‘test model’ upon a similar ‘reference model’ where the correspondence between the test points and the reference points is unknown. For an ICP to reliably give satisfactory results, it is required that:

- the models were pre-aligned to attain the most approximate position.
- the models must go through the segmentation process, which is to remove everything below the teeth area and the posterior part of the model as per the desired measurements.
- the test model should undergo an extra cropping stage to ensure that the boundaries of the model are within the reference model.

2.5 Data processing using AutoIDD

This study was conducted by simulating a dental identification scenario. The user-interface has four sections. It was designed to import digital dental data into “Reference section” (AM) and “Unknown section” (PM) which allows the operator to align the entire datasets to produce results, see Fig. 4. The results will be displayed in the “Results section” and a “viewing section” for inspection of the 3D scans. Thirty AM and 30 PM maxillary dental models (Group A data) were imported into the “reference” and “unknown” sections of the AutoIDD interface respectively. These were aligned for automated comparison (superimposed) and identification of correct matching pairs. This was followed by 30 AM and 30 PM mandibular dental models. The same process was repeated with the 30 sets of IOS (Group B data) and results were obtained. Initially, all the maxillary and mandibular dental models and intra-oral scans were analysed separately to determine the accuracy of AutoIDD software within a specific type of dataset.

To determine the robustness of the software with different scanning methods (laser scanned models and IOS) another 10% of data from each PM dataset (3 from each of 30 3D models/IOS) were selected at random using online random number generator tool [25]. This sample comprised to a total of 12 maxillary and mandibular dental models and IOS (six dental models and six IOS). For the purposes of the experiment, this sample was considered as “test” data. Twelve corresponding/matching 3D models and IOS were retrieved from the respective “reference” data. Additionally, 12 3D models and IOS were also selected at random and included with the 12 corresponding reference data totalling to 24. The 12 “test” data was aligned with 24 “reference” 3D models and IOS.

Once the data is processed, the results were displayed in a tabular form for all the automated alignments. The results sections features a tool bar for data interpretation; such as match score percentage, mean distances and detailed view of the pair-wise alignments. The software generated excel sheets and histograms to export the results. It also generates two colour maps for each 3D pair for qualitative and quantitative evaluation; one for the alignment (qualitative), which shows any discrepancies in the alignment of the arches (in case of a non-match) and one for superimposition (quantitative), to identify the amount of match/non-match areas between

that pair, see Fig. 7. The reference model (AM) is depicted in red while the unknown model (PM) in green for visual differentiation. This software is compatible on a windows laptop/desktop or mac operating system. The statistical analysis of the study groups were assessed using IBM® SPSS Package Version 22 (New York, USA).

3. Results

The results of Group A data showed that the AutoIDD was able to correctly distinguish the matching dental models from the non-matches in all cases. In both the experimental studies (maxillary and mandibular models), all of the matching AM-PM 3D pairs were scored as 100%. This process validates the function of AutoIDD software. The mean distance, standard deviation and RMS (root mean square) of the “correct 3D dental model matches” (maxillary and mandibular) are shown in Table 1. The results of group B data also indicated that the software accurately identified the matching AM-PM IOS from the non-matches in all cases. The match percentage of the maxillary and mandibular intra-oral scans ranged from 64-100% and 81-100%, with a mean percentage of 96.7 and 96.4 respectively. A noteworthy difference in the match percentage between a matching and non-matching pair was observed (Fig. 5). The maximum non-match percentage of the dental models and the scans is shown in Table 1.

Dental Arches	3D scans	AM Data (n)	PM Data (n)	Minimum Match percentage	Maximum Match percentage	Maximum Non-match percentage	Mean Distance (mm) of Correct matches	Mean Distance (mm) of Non-matches	Standard Deviation (mm) of Correct matches	RMS (mm) of Correct matches	RMS of (mm) Non-matches
Maxillary	Dental Models	30	30	-	100	7.7	0.075	0.76	0.006	0.03	1.00
	Intra-oral scans	30	30	64	100	6.0	0.094	0.87	0.031	0.10	1.11
Mandibular	Dental Models	30	30	-	100	8.6	0.076	0.80	0.013	0.06	1.04
	Intra-oral scans	30	30	81	100	6.3	0.093	1.00	0.027	0.11	1.30

Table { SEQ Table * ARABIC } shows the results obtained from the automated comparison of AM-PM 3D datasets; Group A data (Dental Models) and Group B data (Intra-oral scans).

Analysis of the 3D models/IOS displayed a total of 900 automated comparisons and alignments (30 AM x 30 PM) for each data set. Every “Unknown” model (PM) aligned with the given “Reference” (AM) models. This resulted in 30 correct matches based on “best fit alignment” and 870 non-matches. The match percentage of the corresponding AM and PM maxillary and mandibular IOS were indicated in green as shown in Fig. 5 and 6. The results of the “test” data revealed that the AutoIDD clearly differentiated maxillary and mandibular arches and accurately identified correct matching pairs among other non-matches, see Fig. 4. The results (match percentage and mean distances) were consistent with the previous outcomes.

Fig. { SEQ Figure * ARABIC }. shows the user-interface of the AutoIDD software displaying the results of a processed data and 3D dental scan in the viewing section.

The match percentage is computed by the mean distance and the measured error. The mean distance is the distance measured between each reference point on the “unknown” model relative to the corresponding point on the “reference” model. Therefore, pairs with least mean distance between the 3D meshes, have a better alignment and superimposition, and a higher probability of being a “positive identification”. This quantitative measure provides an estimate of the similarity between the two models/scans.

The evaluation of the 3D pairs was performed through the colour maps. Areas with high degree of match in the arches is shown in green with some shades of yellow. Shades of blue and red represent extreme overlaps. Alterations or changes in the dentition (occlusal or any tooth surface) were illustrated in red. For example, a restoration present in the “unknown” scan and not in the “reference” scan is depicted in areas of red, see Fig. 7 (b,d), while an incorrect alignment of a non-match is shown in Fig. 8 a. A colour scale on the superimposition map displays the quantitative differences between the reference points of arches. These coloured maps can be saved in any preferred orientation as JPEG (Joint Photographic Experts Group) images.

Analysis of the total comparisons in each group (maxillary and mandibular 3D models and intra-oral scans) shows the match percentage and histogram of data distribution, see Fig. 9 & 10. Overall, it was possible to identify all the correct matching pairs through the AutoIDD software. The results from the automated comparison and alignment of different types of scans were consistent. There were no scans or matching pairs with similar or close match percentage or mean distance within the processed dataset.

	MAX-PM-01	MAX-PM-02	MAX-PM-03	MAX-PM-04	MAX-PM-05	MAX-PM-06	MAX-PM-07	MAX-PM-08	MAX-PM-09	MAX-PM-10	MAX-PM-11	MAX-PM-12	MAX-PM-13	MAX-PM-14	MAX-PM-15	MAX-PM-16	MAX-PM-17	MAX-PM-18	MAX-PM-19	MAX-PM-20	MAX-PM-21	MAX-PM-22	MAX-PM-23	MAX-PM-24	MAX-PM-25	MAX-PM-26	MAX-PM-27	MAX-PM-28	MAX-PM-29	MAX-PM-30
MAX-AM-01	100	2.6	2.8	2.2	4.8	2.2	3.9	3.1	2.6	5.0	3.4	2.8	2.7	3.3	3.5	2.5	2.3	2.3	2.0	4.9	4.9	2.6	3.1	2.4	2.5	2.1	3.0	1.9	3.2	1.9
MAX-AM-02	2.8	100	2.6	2.7	2.5	4.1	2.9	3.0	3.6	3.0	2.3	2.9	2.0	2.6	3.4	1.8	2.6	3.7	3.0	2.5	2.5	5.2	3.9	2.4	4.9	2.9	3.1	1.9	2.9	2.3
MAX-AM-03	3.3	2.9	100	2.6	3.1	2.5	3.7	2.8	2.7	3.4	2.2	2.7	2.0	2.4	4.2	2.1	2.5	3.1	2.4	2.7	2.9	3.7	3.7	2.5	3.2	2.5	3.0	2.1	2.8	2.3
MAX-AM-04	2.4	2.9	2.7	100	2.3	2.6	2.5	2.5	2.3	2.8	2.2	2.2	1.7	2.5	2.6	1.8	2.7	3.2	2.4	2.2	2.4	3.1	2.7	2.5	2.7	2.6	3.5	2.3	2.9	2.7
MAX-AM-05	6.0	2.7	2.8	1.9	100	2.2	5.0	3.3	2.4	4.6	4.2	4.7	2.4	3.4	4.3	2.8	2.2	2.0	1.9	4.1	4.9	2.7	3.2	2.3	2.5	2.0	2.8	1.8	2.9	1.9
MAX-AM-06	2.1	4.6	2.3	2.3	2.1	100	2.7	2.2	3.7	2.3	2.2	2.5	1.6	2.0	3.3	1.7	3.4	3.7	5.2	1.8	2.0	3.4	2.6	3.8	3.1	4.5	2.7	4.9	3.7	2.8
MAX-AM-07	4.4	3.3	2.9	2.0	5.0	2.3	100	4.0	2.7	4.3	4.0	5.1	2.8	3.1	4.7	2.7	2.0	2.5	2.1	4.3	4.2	3.6	6.0	2.3	4.0	2.0	2.7	2.1	2.8	1.9
MAX-AM-08	3.0	3.2	2.4	2.0	2.8	2.0	3.2	96.1	2.3	3.7	2.6	2.8	2.7	3.5	2.8	2.1	1.9	2.3	1.9	3.1	3.9	3.3	2.1	2.2	1.8	2.6	1.9	2.2	1.8	1.8
MAX-AM-09	2.7	3.1	1.9	2.1	2.6	3.5	3.0	2.2	100	2.6	2.7	3.3	1.7	2.5	4.3	1.8	2.9	2.3	2.3	2.0	2.3	1.8	1.8	3.2	1.7	3.1	1.8	1.6	3.2	2.3
MAX-AM-10	4.7	3.2	2.7	2.1	3.9	2.2	3.6	4.3	2.9	100	3.3	3.4	2.6	2.8	3.2	2.3	2.0	2.3	1.9	3.6	4.7	2.6	2.9	2.4	2.5	2.1	2.4	1.8	2.3	1.9
MAX-AM-11	3.6	2.5	2.6	2.4	3.9	2.3	4.3	3.1	2.4	3.9	100	4.2	2.4	2.7	3.4	2.4	2.0	2.2	1.9	3.4	3.8	3.2	3.5	2.6	2.9	2.1	3.1	1.9	2.4	1.9
MAX-AM-12	3.5	3.9	2.7	2.3	3.9	2.8	4.8	3.5	3.4	4.1	4.2	79.4	2.1	2.9	5.2	2.4	2.3	2.8	2.2	3.0	3.6	4.0	4.7	3.1	4.4	2.4	3.0	1.9	2.9	2.3
MAX-AM-13	2.7	2.1	1.8	1.6	1.9	1.6	2.3	2.8	2.2	2.7	2.4	2.2	99.5	2.1	2.0	2.0	1.4	1.6	1.6	2.6	2.8	1.9	2.1	1.8	1.9	1.5	1.9	1.3	1.7	1.3
MAX-AM-14	3.0	2.7	2.2	2.0	2.9	2.2	2.9	3.5	2.3	2.8	2.7	2.7	2.6	100	3.0	2.9	2.2	2.2	1.9	2.8	3.2	2.5	2.8	2.0	2.6	2.2	2.9	1.8	2.5	1.9
MAX-AM-15	3.3	3.2	2.9	2.4	3.9	3.2	4.7	3.0	3.8	3.3	3.8	4.8	1.9	3.0	100	2.2	2.9	2.8	2.5	2.7	2.9	3.3	4.0	3.0	3.7	2.7	2.4	2.2	3.6	2.6
MAX-AM-16	1.3	1.5	1.9	1.6	1.7	1.1	1.5	1.5	1.2	1.3	1.1	1.1	1.2	1.5	1.3	98.6	1.1	1.5	1.2	1.4	1.6	1.8	1.9	1.1	1.8	1.0	2.1	1.6	1.2	1.1
MAX-AM-17	2.6	3.0	2.0	2.3	2.8	3.4	2.4	2.1	3.1	2.4	2.3	2.5	1.5	2.3	3.0	1.9	100	3.2	2.3	1.8	2.1	2.5	2.6	3.2	2.2	3.4	2.7	2.9	3.7	2.3
MAX-AM-18	2.5	3.8	2.6	2.7	2.1	3.9	2.6	2.3	3.5	2.5	2.2	2.4	1.5	2.1	2.9	1.7	3.1	93.8	2.6	1.9	2.1	3.0	2.6	3.2	2.6	3.9	2.7	2.0	3.2	2.4
MAX-AM-19	2.0	3.9	2.5	2.7	2.0	5.8	2.5	2.1	3.2	2.2	2.2	2.3	1.6	2.0	3.0	1.6	3.0	3.3	100	1.8	1.8	3.0	2.4	2.9	2.7	3.9	2.4	4.1	3.3	4.8
MAX-AM-20	5.6	2.8	2.2	1.8	4.2	2.1	3.2	3.6	2.6	4.8	3.6	3.6	3.3	3.2	3.1	2.3	2.0	2.1	2.1	100	4.0	2.7	3.7	2.3	4.2	1.9	3.2	1.7	2.5	1.7
MAX-AM-21	4.9	2.8	2.1	2.0	4.2	1.9	3.1	4.3	2.2	4.7	3.7	3.4	3.0	3.7	3.1	2.9	1.8	2.2	1.8	4.0	100	3.2	3.8	2.2	3.2	1.8	3.4	1.8	2.3	1.9
MAX-AM-22	2.9	4.7	3.0	2.9	3.0	3.1	3.6	3.5	3.1	3.1	2.8	3.6	2.0	2.9	4.4	2.1	2.5	3.3	2.6	2.7	2.9	98.9	4.8	2.6	4.7	2.6	4.1	2.2	2.9	2.6
MAX-AM-23	3.6	3.5	2.9	2.5	3.8	2.9	5.3	3.4	3.6	3.5	3.2	4.3	2.4	3.0	5.5	2.2	2.4	2.8	2.4	3.4	3.4	4.8	89.7	3.1	4.6	2.5	2.7	2.0	3.2	2.3
MAX-AM-24	2.5	3.9	2.5	2.6	2.5	4.1	2.7	2.4	3.4	2.5	2.5	2.8	1.8	2.2	3.0	1.8	3.1	3.5	2.3	2.1	2.2	3.1	3.1	100	3.1	5.1	3.1	2.1	3.2	2.2
MAX-AM-25	3.2	5.1	2.5	2.4	2.9	2.8	3.7	3.1	3.2	2.8	2.5	3.7	2.0	2.8	4.1	2.0	2.2	3.0	2.3	2.9	2.9	5.0	4.4	2.9	91.1	2.3	3.4	2.2	2.7	2.4
MAX-AM-26	2.3	3.8	2.3	2.4	2.3	4.7	2.6	2.1	3.4	2.3	2.1	2.3	1.7	2.3	2.9	1.9	3.2	4.2	2.3	1.9	2.1	3.2	2.9	4.4	2.6	91.6	3.0	3.3	3.3	3.9
MAX-AM-27	3.7	3.5	2.4	2.8	3.3	2.6	3.4	3.2	2.7	3.0	3.0	3.0	2.3	3.3	3.6	2.4	2.5	2.7	2.4	3.1	3.2	4.3	2.6	3.3	2.6	100	2.3	3.4	2.7	2.7
MAX-AM-28	2.5	3.8	2.4	2.7	2.3	6.1	2.7	2.2	3.2	2.2	2.5	2.8	1.6	2.4	3.1	2.1	4.0	3.5	3.8	2.0	2.3	2.9	2.6	3.4	2.7	4.3	3.1	64.6	3.3	3.0
MAX-AM-29	2.7	2.9	2.3	2.4	2.6	3.5	2.6	2.3	3.1	2.5	2.6	2.7	1.5	2.2	2.9	1.6	2.8	2.5	2.9	2.0	2.2	2.2	2.3	3.0	1.9	3.1	2.2	2.3	100	2.6
MAX-AM-30	2.2	4.7	2.5	2.8	2.4	3.3	2.7	2.3	3.4	2.5	2.1	2.5	1.5	2.2	3.4	1.9	4.0	5.0	4.9	1.8	2.1	3.5	2.6	3.5	3.3	4.9	2.9	5.1	3.5	100

Fig. { SEQ Figure * ARABIC }. Match percentages of AM – PM maxillary intra-oral scans. The correct matching pairs

→

	MAND-PM-01	MAND-PM-02	MAND-PM-03	MAND-PM-04	MAND-PM-05	MAND-PM-06	MAND-PM-07	MAND-PM-08	MAND-PM-09	MAND-PM-10	MAND-PM-11	MAND-PM-12	MAND-PM-13	MAND-PM-14	MAND-PM-15	MAND-PM-16	MAND-PM-17	MAND-PM-18	MAND-PM-19	MAND-PM-20	MAND-PM-21	MAND-PM-22	MAND-PM-23	MAND-PM-24	MAND-PM-25	MAND-PM-26	MAND-PM-27	MAND-PM-28	MAND-PM-29	MAND-PM-30	
MAND-AM-01	97.9	2.5	2.0	2.9	1.7	1.4	2.5	2.8	1.8	3.7	2.8	1.8	2.1	1.9	3.7	2.4	1.9	1.8	1.8	3.2	3.2	2.6	1.6	1.5	2.9	2.6	3.2	1.9	2.9	2.5	
MAND-AM-02	2.6	88.5	2.3	2.4	1.4	3.7	2.7	2.5	4.3	2.2	2.2	3.6	1.7	2.4	3.4	3.1	3.2	3.1	2.8	2.1	2.5	3.7	5.3	2.5	3.4	5.0	3.3	1.9	3.9	6.0	
MAND-AM-03	2.4	2.2	99.9	2.8	2.0	2.0	3.4	1.9	2.1	3.2	2.3	2.8	1.3	1.7	3.0	2.4	1.5	1.7	2.5	2.6	2.2	3.4	2.6	2.2	3.6	1.9	3.0	2.1	2.1	2.3	
MAND-AM-04	2.1	2.2	2.2	91.0	2.0	1.6	2.2	2.1	2.0	3.5	3.3	2.7	2.3	2.2	3.5	2.7	1.7	2.0	2.6	2.7	2.9	2.5	1.8	1.4	4.2	2.2	2.4	1.4	2.4	2.2	
MAND-AM-05	1.5	2.4	2.3	1.4	87.3	2.3	3.0	1.8	3.1	1.7	2.8	3.1	1.2	1.8	1.8	3.2	2.5	1.9	2.4	1.2	3.3	2.2	2.6	2.3	2.7	2.2	1.7	1.1	2.2	2.4	
MAND-AM-06	2.3	5.0	2.0	2.2	1.7	100	2.4	2.3	2.8	2.1	1.9	2.3	1.8	1.9	2.7	2.5	2.4	2.8	2.1	1.9	2.2	3.3	3.2	1.7	2.8	4.2	3.0	1.8	3.0	5.6	
MAND-AM-07	3.2	2.7	3.3	2.6	0.3	2.4	100	2.0	3.8	4.7	3.3	3.6	1.3	1.8	5.0	2.4	2.3	1.9	3.2	4.4	2.9	4.1	3.3	3.3	4.9	2.0	4.0	1.8	2.7	2.7	
MAND-AM-08	1.9	2.4	1.5	2.6	1.2	1.8	1.6	100	2.3	1.8	2.3	2.1	3.6	2.5	2.3	3.6	2.4	2.4	1.7	1.9	2.7	1.8	1.6	1.5	2.0	3.0	2.1	1.8	3.2	2.2	
MAND-AM-09	1.2	2.4	2.2	1.0	1.1	3.0	2.3	1.3	100	1.3	1.2	2.2	1.5	2.0	1.6	1.5	3.1	2.2	2.5	1.4	1.3	2.1	2.5	2.6	1.9	2.6	1.7	1.3	2.3	2.4	
MAND-AM-10	3.9	2.1	3.0	3.0	1.1	1.5	3.9	1.8	2.7	100	3.2	2.1	1.3	1.5	3.5	2.9	1.7	1.6	2.9	4.0	3.9	2.8	2.1	1.8	3.6	1.9	3.0	1.5	2.2	2.2	
MAND-AM-11	1.7	2.2	2.0	2.1	1.4	1.6	2.4	2.1	2.3	2.0	100	2.1	1.8	1.7	1.9	2.8	1.9	1.7	2.2	2.2	3.8	2.2	2.0	1.8	2.5	2.0	1.7	1.2	2.0	2.0	
MAND-AM-12	2.5	4.2	2.7	3.1	0.3	3.1	3.5	2.6	3.5	2.5	2.4	100	1.7	2.6	3.7	3.8	2.4	2.6	2.7	2.7	2.6	3.3	3.6	2.9	3.6	3.5	3.2	1.4	4.0	3.6	
MAND-AM-13	1.5	1.4	1.2	2.3	1.1	1.4	1.2	2.7	1.7	1.9	1.2	1.2	89.3	1.9	2.2	1.9	1.4	1.6	1.2	1.3	1.0	1.3	1.4	1.3	1.2	2.0	1.6	2.9	1.8	1.4	
MAND-AM-14	2.0	2.8	1.8	2.4	1.2	1.8	1.6	100	2.3	2.4	1.6	1.9	2.4	2.2	98.4	2.1	2.4	2.3	2.1	1.7	2.0	2.3	2.4	1.7	2.4	2.7	2.2	2.6	3.8	2.5	
MAND-AM-15	3.2	3.1	2.7	2.6	1.4	2.1	4.9	2.1	4.4	3.4	3.1	3.2	1.4	2.2	99.2	2.4	2.2	3.0	3.1	3.2	3.6	3.2	2.2	4.3	2.6	2.7	1.4	3.2	3.1		
MAND-AM-16	2.4	2.9	2.0	2.2	0.8	2.3	1.0	3.5	0.9	2.8	2.6	2.1	2.4	2.4	2.7	97.6	2.6	3.0	2.6	1.5	3.6	2.6	1.8	2.3	2.0	2.8	2.9	2.6	2.2	3.0	2.8
MAND-AM-17	2.5	3.7	2.5	3.2	2.4	2.3	2.4	2.3	2.9	2.4	2.3	2.9	2.4	2.3	2.9	2.4	2.3	2.9	2.4	2.3	2.9	2.4	2.3	2.9	2.4	2.3	2.9	2.4	2.3	2.9	2.4
MAND-AM-18	1.6	2.3	1.5	1.6	1.9	1.6	2.7	2.5	1.4	1.7	2.2	2.4	2.3	1.8	3.1	2.4	94.7	1.8	1.4	1.9	1.8	1.8	1.6	1.8	1.6	1.8	3.0	1.7	1.3	2.8	2.4
MAND-AM-19	2.7	2.9	3.0	2.7	0.2	2.7	3.3	1.8	3.2	3.1	2.6	2.8	1.5	1.7	3.1	2.8	2.0	1.9	100	2.8	2.6	3.3	2.9	2.7	3.7	1.9	3.1	1.1	2.4	2.6	
MAND-AM-20	3.4	2.5	2.7	2.4	2.9	2.1	4.4	2.0	3.1	4.9	3.6	3.1	1.3	1.9	3.6	2.6	2.2	1.8	2.4	100	3.5	2.9	2.7	2.4	3.4	2.2	3.2	1.8	2.6	2.7	
MAND-AM-21	2.9	2.4	1.9	3.1	0.9	1.7	2.3	2.6	2.7	3.4	3.9	2.1	2.7	1.8	3.0	2.8	2.0	1.9	2.4	2.9	100	2.3	1.9	1.5	2.5	1.7	2.4	1.5	2.7	2.4	
MAND-AM-22	2.7	3.2	2.6	3.2	1.5	2.3	2.4	2.3	2.9	2.5	2.3	3.7	2.6	2.3	3.7	3.6	2.0	2.8	2.2	2.6	95.6	2.7	3.8	5.3	3.0	3.2	3.3	3.9	3.9	3.9	
MAND-AM-23	2.6	5.5	2.5	1.6	1.5	2.3	3.2	2.6	3.0	2.5	2.8	3.3	1.7	2.3	4.5	2.5	2.4	2.4	2.3	2.9	96.3	2.9	96.3	2.9	2.9	3.8	3.2	2.1	3.1	5.6	
MAND-AM-24	3.1	2.6	2.7	2.9	0.2	1.8	4.0	2.2	2.3	3.2	2.5	2.6	1.6	1.9	3.9	2.4	2.0	2.0	2.4	2.9	2.5	3.9	3.1	86.1	3.9	2.6	3.8	2.2	2.5	3.4	
MAND-AM-25	2.8	3.4	3.3	3.0	2.5	2.8	4.0	2.0	2.9	3.3	2.8	3.4	1.4	1.9	3.9	2.8	2.1	2.0	3.2	2.9	2.6	5.3	3.7	2.9	99.8	2.9	3.5	2.2	2.8	3.4	
MAND-AM-26	2.6	5.5	2.2	2.9	0.4	2.7	2.3	3.5	2.7	3.4	2.2	2.2	2.1	2.2	2.5	3.0	3.0	3.5	3.3	2.2	2.0	3.1	3.8	2.1	2.9	81.9	3.0	2.5	3.8	3.4	
MAND-AM-27	3.9	3.2	3.0	2.9	1.3	2.7	3.7	2.0	2.7	2.5	2.2	2.5	3.1	1.4	2.1	3.4	2.4	2.3	2.2	2.9	3.4	2.6	3.6	2.6	3.8	2.7	100	2.0	2.1	3.6	
MAND-AM-28	3.2	1.9	3.2	3.2	1.2	1.3	2.7	1.8	1.8	3.7	2.7	2.7	2.2	2.5	1.8	1.9	1.8	1.9	1.8	1.9	1.8	1.9	1.8	1.9	1.8	1.9	1.8	1.9	1.8	1.9	
MAND-AM-29	2.5	3.1	1.9	2.6	1.3	2.3	2.5	2.1	3.1	3.6	2.0	2.3	3.1	3.0	3.6	2.9	3.4	2.7	2.4	2.2	2.1	2.6	2.5	2.1	1.9	2.6	2.4	1.9	100	3.3	
MAND-AM-30	2.5	6.3	2.6	2.3	2.5	5.7	2.8	2.2	3.8	2.3	2.4	3.6	1.5	2.3	3.4	3.0	3.9	3.1	2.6	2.1	2.5	3.7	5.9	3.0	3.5	4.7	3.3	1.9	3.6	100

Fig. { SEQ Figure * ARABIC }. Example of an incorrect alignment (a) and superimposition (b) of a non-matching AM (Red)

Fig. 9. Histogram of the distribution of percentage of matches and non-matches of maxillary IOS.

Fig. 10. Histogram of the distribution of percentage of matches and non-matches of mandibular IOS.

4. Discussion

This study demonstrates on the feasibility of using 3D dental models and IOS in the new automated system for accurate dental identification. There is a shift towards digital storage of dental records, facilitated by affordable 3D scanners which can digitise the patient's dental casts [17]. Through the increasing use of 3D scanners and chairside intraoral scanners [12-16], 3D dental models may be considered as a useful source of AM information. All the scientific primary methods of human identification (fingerprint, DNA analysis, and dental) involve the comparison of the AM data to PM evidence to establish a positive identification [2]. Furthermore, the dental characteristics; their alignment and orientation within the arch, tooth shape and dental treatment interventions were considered for a dentition to be unique [26].

Using registration techniques, the likelihood of finding matching dentitions in a given population were conducted in 2D [27-29]. In the process of investigating the uniqueness of the human dentition, few attempted matching studies on 3D scanned dentitions [30,31], while orthodontically treated samples were evaluated in 3D using semi-automatic methods [32,33] which were all landmark based. All of these studies examined only the upper and lower anterior teeth. A pilot study [34] analysed the incisal edges (2 mm) of the six anterior teeth of post-orthodontic treated dentitions using 3D software packages for identifying matching set of dentitions. Though, the study concluded that the human anterior dentition is unique and encouraged 3D approach, it required manual segmentation and lacked complete automation. Another study [35] presented an algorithm for automatic matching using scanned mandibular 3D models. The study was based on feature extraction and description points from the models and concluded that the accuracy was highest with manual-segmented models than auto-segmented PM models.

The present study is unique from the above studies. The main purpose of the AutoIDD software is to align and identify matching dental patterns from the reference data. The entire scan area is considered in the registration process mainly focussing on the dental components. All the previous studies used existing softwares which have some limitations. The major advantage of AutoIDD software is that it was designed and developed to be fully automated and no manual intervention is required. The target scan in a large dataset was efficiently identified without any false positives. It also eliminates the need for surface registration and manual segmentation.

Though various studies proposed similar methods in the recent past, there are no reported studies which presented a fully automated system to assist forensic experts with dental identification. A clear distinction between match and non-match, robustness with different scan types and overall efficiency and accuracy in achieving the results makes this study novel. The software assigns a match percentage to any given comparison and with multiple different data and scenarios being executed.

Gibelli et al. also proposed similar procedures of 3D-3D superimposition, one that analysed anatomical differences between dental elements [36] and another verifying the uniqueness of 3D models of the palatal rugae [37]. The superimposition was based on the surface registration, where the software used calculated the least point-to-point distance between the two superimposed surfaces, expressed as RMS. As the AutoIDD software considers the centre of each triangle in the mesh/scan surface, all the points on the test scan are compared to the nearest corresponding points on the reference scan. The mean of all the point-to-point (minimum) distance is referred to as 'mean distance' in the results. There are no specific landmarks involved in the process.

It was observed that the orthodontic treatment increased the similarity and reduced the uniqueness of the human anterior dentition between different patients. These similarities may cause the dentitions to be less distinguishable than before treatment [33]. The study data used in group A were all 3D dental models of post-orthodontic treated patients. The results of the experimental study indicate that the software has correctly distinguished the matching dental models from the non-matches and all the models were found to be a 100% match. This also shows that there were no changes detected in the AM-PM matching pairs. No false positives were observed. This part of the study validated the function of AutoIDD. The group B data consisted of random individuals from general population relevant only to the locality. In a forensic context, though equal number of corresponding or matching 3D reference scans were considered during the automated process, there was only one correct AM match in the reference data for every PM case. The remaining 29 were non-matches. The AutoIDD was found to be consistent in accuracy; capable of identifying “correct matches” from “non-matches” by assigning a high match percentage for correct match and a very low for a non-match.

There may be more than one ante-mortem dental record, compared to one post-mortem dental record, if a number of individuals are missing and the authorities suspect that the unidentified deceased could be any one of the indicated missing persons. Therefore, the automated process was subjected to test the performance with an increased number of AM non-matches/cases with a “test” data. The software demonstrated its robustness in efficiently performing with different types of 3D scans, see Fig. 4.

The important outcome of the AutoIDD software is that, for a correct match the match percentage is higher than other scores and it narrows down to the correct AM correspondence (indicated in green), see Fig. 5, 6. Hence, low or high match percentage is not only the point, but needs to be discriminatory. For example when a 64% is assigned to a PM-1 and AM-1 3D pair and the next nearest match percentage for PM-1 and AM-2 is approximately 3%, and no other pair has a higher match percentage, it indicates that PM-1 and AM-1 is the best matching pair for that PM model. All these numbers are content dependant; variations in the dentitions, see Fig. 7. Similarly, 100% means that within the precision of the software and within the boundaries of its margin for error, it is able to determine that it is an exact match, i.e. best fit match.

The intraoral scan data had some dental alterations (therapeutic and non-therapeutic) during the scan interval. Many of the changes a tooth can undergo are by their nature irreversible. These discrepancies are significant during the comparison process and must be evaluated by the forensic expert as to whether they are explainable or unexplainable [2]. An ‘explainable discrepancy’ is one where an unrestored tooth/teeth surface(s) in AM model is found to be restored in the PM model. While a tooth/teeth present in the PM model and missing in the AM model is considered as an ‘unexplainable discrepancy’. The sensitivity towards any changes in the dentition can be appreciated with the help of colour maps generated for every 3D pair. It was also noticed that the influence of soft tissues on the overall result/score was negligible when compared to the dental component.

The evidence presented in this paper strongly suggests that for a given comparison, the correct match will be the one with the highest match percentage and lowest mean value. It was observed that the accuracy of the alignment process and the match percentage increased when there was greater similarity between the dentitions being compared. Where changes in the dentition had occurred between scans, the match percentage was reduced. In addition, the whole arch presentation may enhance the overall reliability of the identification process. The

advantage of 3D format is the dimensional stability; correct size proportions, which is a limitation for the radiographs.

With the rapidly advancing 3D technologies, computerized tomography data and 3D scan data are considered more reliable AM records than written dental charts. The availability of AM 3D data can allow for an accurate digital comparison with the PM 3D data which can be applied in single identification cases. The process of 3D comparison may potentially be considered in the Disaster Victim Identification techniques in future with the increased accessibility and use of 3D datasets [38]. Henceforward, the dentists are recommended to treat diagnostic casts as part of the dental record and preserve them for several years following the completion of the dental treatment depending on the residence countries dental association's regulations and guidelines. This study supports the storage of dental casts as a component of AM records. The forensic odontologists are also encouraged to obtain PM dental impressions whenever possible which may facilitate a process of automation in forensic identification.

This is the first fully automated system which is intended to assist forensic experts in the process of dental identification. The performance of the AutoIDD has so far been tested only with full arch dental scans. Through this study it is established that this automated identification system is consistent with the construct it is supposed to be measuring, which is a "proof of concept" and very encouraging at the early stages of development. Moreover, AutoIDD may also be used in the digital retrieval of patient dental records from digital dental databases and in dental practices. Further research is required in the investigation of partial remains, pre- and post-orthodontic samples and single tooth PM 3D models.

5. Conclusion

In summary, AutoIDD was able to demonstrate the identification of correct matches with a match percentage that clearly differentiates the matches from non-matches. This software also enables recognition of the changes in the human dentition, such as restorations and missing teeth. The match percentage is a combination of the best fit alignment and the morphological changes identified by the software. A low match percentage may also infer that there has been significant changes to that individual's dentition usually due to dental intervention. These changes are also visible on the superimposition heat maps produced by the software.

This study presents a novel method for automated human dental identification using 3D scans to record digital data, the innovative utility of python software, match percentage algorithm and image superimposition.

Conflict of interest: The authors declare that they have no conflict of interest.

Funding: None declared.

6. References

1. J. Hinchliffe, Forensic odontology, Part 1. Dental identification, Br. Dent. J. 210(5) (2011) 219-24.
2. R. Carabott, Dental human identification, In: C. Adams, R. Carabott, S. Evans (Eds.), Forensic Odontology: An Essential Guide, 1st edn. John Wiley & Sons, Ltd, West Sussex, UK, 2014, pp. 65-110.
3. L. Lorton, M. Rethman, R. Friedman, The computer-Assisted Postmortem Identification (CAPMI) System: a computer-based identification program, J. Forensic Sci. 33(4) (1988) 977-984.
4. J. McGivney, WinID3-Dental Identification System, 2006. <http://www.winid.com> (accessed November 20, 2019).

5. J.G. Clement, V. Winship, J. Ceddia, S. Al-Amad, A. Morales, A.J. Hill, New software for computer-assisted dental-data matching in disaster victim identification and long-term missing persons investigations: “DAVID Web”, *Forensic Sci. Int.* 159(1) (2006) 24–29.
6. L. Andersen Torpet, DVI system international: software assisting in the Thai tsunami victim identification process, *J. Forensic Odontostomatol.* 23 (2005) 19–25.
7. A.K. Jain, H. Chen, Matching of dental x-ray images for human identification, *Pattern Recognit.* 37 (7) (2004) 1519–1532.
8. H. Chen, A.K. Jain, Dental biometrics: alignment and matching of dental radiographs, *IEEE Trans. Pattern Anal. Mach. Intell.* 8 (2005) 1319–1326.
9. G. Fahmy, H. Chen, O. Nomir, R. Howell, M. Abdel-Mottaleb, A.K. Jain, H. H. Ammar, J. Zhou, D. Nassar, E. Haj-Said, Toward an automated dental identification system, *J. Electron. Imaging*, 14 (4) (2005) 1–13.
10. O. Nomir, M. Abdel-Mottaleb, Fusion of matching algorithms for human identification using dental X-ray radiographs, *IEEE Trans. Inf. Forensics Secur.* 3 (2) (2008) 223–233.
11. P.-L. Lin, Y.-H. Lai, P.-W. Huang, Dental biometrics: human identification based on teeth and dental works in bitewing radiographs, *Pattern Recognit.* 45 (3) (2012) 934–946.
12. D.J. Fasbinder, Digital dentistry: Innovation for restorative treatment, *Compend. Contin. Educ. Dent.* 31(2010) 2-11.
13. D. Naidu, T.J. Freer, Validity, reliability, and reproducibility of the iOC intraoral scanner: a comparison of tooth widths and Bolton ratios, *Am. J. Orthod. Dentofac. Orthop.* 144 (2013) 304–310.
14. S. Ting-Shu, S. Jian, Intraoral Digital Impression Technique: A Review, *J. Prosthodont.* 24(4) (2015) 313–21. <https://doi.org/10.1111/jopr.12218>.
15. E.V. Chalmers, G.T. McIntyre, W. Wang, T. Gillgrass, C.B. Martin, P.A. Mossey, Intraoral 3D Scanning or Dental Impressions for the Assessment of Dental Arch Relationships in Cleft Care: Which is Superior? *Cleft Palate Craniofac. J.* 53(5) (2016) 568-77. <https://doi.org/10.1597/15-036>.
16. C. Wesemann, J. Muallah, J. Mah, A. Bumann, Accuracy and efficiency of full-arch digitalization and 3D printing: A comparison between desktop model scanners, an intraoral scanner, a CBCT model scan, and stereolithographic 3D printing. *Quintessence Int.* 48(1) (2017) 41-50. <https://doi.org/10.3290/j.qi.a37130>.
17. Martin CB, Chalmers EV, McIntyre GT, Cochrane H, Mossey PA. Orthodontic scanners: what’s available? *J. Orthod.* 42 (2015) 136–143.
18. B. Bustos, D.A. Keim, T. Schreck, D. Vranic, An experimental comparison of feature based 3D retrieval methods, In: 2nd International Symposium on 3D Data Processing, Visualization and Transmission; Sep. 6-9; Thessaloniki, Greece, IEEE Publication, (2004) 215–222.
19. M. Chaouch, A. Verroust-Blondet, Alignment of 3D models, *Graphical Models*, 71(2) (2009) 63-76.
20. C. He, C. Wang, A Survey on Segmentation of 3D Models, *Wireless Pers. Commun.* 102(4) (2018) 3835-42.
21. T. Kronfeld, D. Brunner, G. Brunnert, Snake-based segmentation of teeth from virtual dental casts, *Comput. Aided Des. Appl.* 7(2) (2010) 221–233.
22. T. Kondo, S.H. Ong, K.W.C. Foong, Tooth segmentation of dental study models using range images. *IEEE Trans. Med. Imaging*, 23(3) (2004) 350–362.
23. Python software; Available from <https://www.python.org/> (accessed October 12, 2018).
24. P.J. Besl, N.D. McKay, A method for registration of 3-D shapes, *IEEE Trans. Anal. Mach. Intell.* 14(2) (1992) 239-256.
25. True Random number generator. <https://www.random.org/> (accessed July 14, 2019).
26. A. Franco, G. Willems, P.H.C. Souza, G.E. Bekkering, P. Thevissen, The uniqueness of the human dentition as forensic evidence: a systematic review on the technological methodology, *Int. J. Legal Med.* 129(6) (2015) 1277-83. <https://doi.org/10.1007/s00414-014-1109-7>.
27. J.A. Kieser, V. Bernal, J.N. Waddell, S. Raju, The uniqueness of the human anterior dentition: a geometric morphometric analysis, *J. Forensic Sci.* 52 (3) (2007) 671-677. <https://doi.org/10.1111/j.1556-4029.2007.00403.x>
28. M.A. Sheets, P.J. Bush, H.D. Bush, Statistical evidence for the similarity of the human dentition, *J. Forensic Sci.* 56 (1) (2011) 118-23. <https://doi.org/10.1111/j.1556-4029.2010.01531.x>.
29. H.D. Sheets, P.J. Bush, C. Brzozowski, L.A. Nawrocki, P. Ho, M.A. Bush, Dental shape match rates in selected and orthodontically treated populations in New York State: a 2-dimensional study, *J. Forensic Sci.* 56 (3) (2011) 621-6. <https://doi.org/10.1111/j.1556-4029.2011.01731.x>.
30. M.A. Bush, P.J. Bush, H.D. Sheets, Similarity and match rates of the human dentition in three dimensions: relevance to bitemark analysis, *Int. J. Legal Med.* 125 (2011) 779–784, <http://dx.doi.org/10.1007/s00414-010-0507-8>.
31. H.D. Sheets, P.J. Bush, M.A. Bush, Patterns of variation and match rates of the anterior biting dentition: characteristics of a database of 3-D scanned dentitions, *J. Forensic Sci.* 58 (2013) 60-68. <https://doi.org/10.1111/j.1556-4029.2012.02293.x>

32. A. Franco, G. Willems, P.H.C. Souza, O.M. Tanaka, W. Coucke, P. Thevissen, Three-dimensional analysis of the uniqueness of the anterior dentition in orthodontically treated patients and twins, *Forensic Sci. Int.* 273 (2017) 80–87.
33. A.E.C. Dyke, S. Cunningham, N. Hunt, C. Ruff, A comparative study to investigate the effect of orthodontic treatment on the uniqueness of the human anterior dentition, *Forensic Sci. Int.* 289 (2018) 368–373. <https://doi.org/10.1016/j.forsciint.2018.06.008>.
34. G.T.F. Chong, A.H. Forgie, A Pilot Study to Analyze the Uniqueness of Anterior Teeth Using a Novel Three-Dimensional Approach, *J. Forensic Identif.* 67(3) (2017) 361–378.
35. X. Zhong, D. Yu, Y.S. Wong, T. Sim, W.F. Lu, K.W.C. Foong, Ho-L. Cheng, 3D dental biometrics: Alignment and matching of dental casts for human identification, *Comput. Ind.* 64(9) (2013) 1355–1370.
36. D. Gibelli, D. De Angelis, F. Riboli, C. Dolci, C. Cattaneo, C. Sforza, Quantification of odontological differences of the upper first and second molar by 3D-3D superimposition: a novel method to assess anatomical matches, *Forensic Sci. Med. Pathol.* 15(4) (2019) 570–573.
37. D. Gibelli, D. De Angelis, V. Pucciarelli, F. Riboli, V.F. Ferrario, C. Dolci, C. Sforza, C. Cattaneo, Application of 3D models of palatal rugae to personal identification: hints at identification from 3D-3D superimposition techniques, *Int. J. Legal Med.* 132 (2018) 1241–1245. <https://doi.org/10.1007/s00414-017-1744-x>
38. A. Forrest, Forensic odontology in DVI: current practice and recent advances, *Forensic Sci. Res.* 4 (2019) 316–330. <https://doi.org/10.1080/20961790.2019.1678710>.