University of Dundee

Introduction of a Measurement Setup to Monitor the Pressure Applied During Handheld Ultrasound Elastography

Skerl, Katrin; Eichhorn, Bernd; Poltorjanoks, Romans; Cochran, Sandy; Evans, Andrew

Published in:
Ultrasound in Medicine and Biology

DOI:
[10.1016/j.ultrasmedbio.2020.04.017](https://doi.org/10.1016/j.ultrasmedbio.2020.04.017)

Publication date:
2020

Licence:
CC BY-NC-ND

Document Version
Peer reviewed version

[Link to publication in Discovery Research Portal](#)

Citation for published version (APA):

Skerl, K., Eichhorn, B., Poltorjanoks, R., Cochran, S., & Evans, A. (2020). Introduction of a Measurement Setup to Monitor the Pressure Applied During Handheld Ultrasound Elastography. *Ultrasound in Medicine and Biology*, 46(9), 2556-2559. <https://doi.org/10.1016/j.ultrasmedbio.2020.04.017>

General rights

Copyright and moral rights for the publications made accessible in Discovery Research Portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from Discovery Research Portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain.
- You may freely distribute the URL identifying the publication in the public portal.

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Introduction of a measurement set-up to monitor the pressure applied during hand-held ultrasound elastography

Katrin Skerl^{a,1,*}, Bernd Eichhorn^c, Romans Poltorjanoks^d, Sandy Cochran^d, Andrew Evans^a

^a*University of Dundee, Dundee, Scotland, UK*

^b*Furtwangen University, Villingen-Schwenningen, Germany*

^c*Schuler-Konstruktionen GmbH, Nuremberg, Germany*

^d*University of Glasgow, Glasgow, Scotland, UK*

Abstract

Shear wave elastography may produce misleadingly high values if too much pressure is applied during the imaging process. However, in clinical routine there is presently no way to monitor the pressure applied during the measurements. In this work we introduce a novel measurement set-up which can directly be attached to an ultrasonic imaging transducer and allows to observe the applied pressure in real time. The introduced set-up supports free-hand imaging according to the clinical standard. We tested the set-up by carrying out SWE under varying pressures on *ex vivo* animal tissue. The SWE values increased with pressure as was expected. Thus, the introduced set-up is a possible solution to measure the applied pressure in real-time.

Keywords: Ultrasound, Pressure, Elastography, Shear Wave Elastography, Cancer, Breast Cancer, Phantoms

*Corresponding Author: Katrin Skerl, Faculty of medical and mechanical engineering, Furtwangen University, Jakob-Kienzle Str. 16, 78054 Villingen-Schwenningen, Germany; Email: katrin.skerl@hs-furtwangen.de, Phone: +49 7720 307 4506

1 **Introduction**

2 Supersonic shear wave elastography (SWE) is an ultrasound (US) imaging
3 technique used to visualise and measure the elasticity of tissue. Already the
4 first clinical SWE study showed that applied pressure influences the SWE
5 measurements and recommended avoiding the application of pressure during
6 the procedure (Berg et al., 2012). However, SWE imaging is carried out
7 manually and it is thus, difficult to completely avoid pressure applied by the
8 clinician. Even if the clinician deliberately relieves her or his pressure, the
9 weight of the probe may still cause some pressure to be applied. Hence, it is
10 reasonable to state that application of a constant and reproducible pressure
11 is impossible to achieve in clinical practice.

12 Some previous studies showed the influence of applied pressure on mea-
13 surement of SWE values for breast tissue (Barr and Zhang, 2012; Wojcinski
14 et al., 2013; Bernal et al., 2016; Sayed et al., 2014). Their set-up is not appli-
15 cable in clinically routine SWE imaging, where the patient is in the supine
16 position. In response to the needs, capabilities and limitations of previous
17 work, here we introduce a novel measurement device, which can be attached
18 around an US SWE transducer to monitor the applied pressure in real time
19 with as little impact as possible onto the clinical protocol.

20 **Materials and Methods**

21 The novel measurement device has two requirements to aid clinical rou-
22 tine. First, it should be easily attached around the imaging probe to allow
23 SWE measurements to be made in its presence and absence to avoid bias in

24 clinical practice. Second, the attachment and removal processes should take
25 a maximum of a few minutes to allow application in routine clinical practice.

26 Three observers (2 radiologists, 1 radiographer, all with at least 5 years'
27 experience in breast sonography and 3 years' experience in SWE imaging)
28 were asked to apply pressure with the SWE probe on a commercial phantom
29 (CIRS, Model 059, Norfolk, VI, USA) similar to what they would apply
30 in clinical routine. The phantom was placed on a calibrated scale and the
31 increase in weight was monitored. The applied pressure p was calculated
32 over the transducer's surface (61 mm x 8 mm, $A = 488 \text{ mm}^2$). The applied
33 p differed by at least 0.2 N / 0.4 kPa for the three evaluated observers.
34 Therefore, the sensor should enable an accuracy of 0.1 N / 0.2 kPa. The
35 upper limit of the measurement range was estimated after discussion with an
36 expert radiologist with more than 20 years' experience in US breast imaging
37 and was, accordingly, set to 10 N or 20 kPa.

38 The measurement device is realised as a double shell around the SWE
39 probe **to avoid any damage of the probe. The pressure is** measured
40 **between these shells. The inner shell (Fig. 1a) is attached directly**
41 **to the SWE probe, while the outer shell (Fig. 1b) is moveable**
42 **relative to the inner shell via a sliding system. A spring pushes the**
43 **outer shell upwards onto the pressure sensors, which are attached**
44 **to the inner shell. Thus, the device introduced here measures a**
45 **reduction rather than an increase in applied pressure to allow the**
46 **measurement of unladen pressure.**

47 The shells were created by 3D printing using acrylonitrile butadiene
48 styrene (ABS) plastic. This allowed freedom in the design while keeping

49 production costs low. Four springs press the outer shell upwards with a force
 50 of 10 N, i.e. the maximum of the measurement range defined. Two strain-
 51 gauge sensors (FSR402, Interlink Electronics, Los Angeles, CA, USA) are
 52 positioned on each side of a step leading around the inner shell and convert
 53 the applied force into a decrease in electrical resistance and thus, a decrease
 54 in voltage drop. The sensors are positioned out of alignment with the centre
 55 to allow the measurement of any drift if the pressure is not applied exactly
 56 vertically. Careful handling of the sensors was required as they broke easily.
 57 The outer shell presses against the pressure sensors. When the SWE probe
 58 touches the tissue the pressure on the sensors decreases and a lower signal is
 59 produced.

60 Measurements with and without the SWE probe included within the
 61 shells were calibrated to validate the correlation between the applied pres-
 62 sure and the signal outputs. The calibration was performed with the device
 63 placed directly onto a calibrated scale. The measurements on the *ex vivo*
 64 samples were performed according to routine clinical practice but with a cal-
 65 ibrated scale underneath. The scale used for the measurements was identical
 66 with the one used in the calibration process. The preloaded pressure was
 67 increased from 0 g / 0 N / 0 kPa to 1 kg / 10 N / 20 kPa with increments of
 68 100 g / 1 N / 2 kPa. For each preload, five measurements were recorded and
 69 averaged for evaluation.

70 ***Ex vivo*** samples including chicken breast, porcine belly, boiling beef and
 71 bovine udder were investigated. The *ex vivo* tissues were obtained through
 72 a local slaughterhouse and no additional harm was applied to any animal.
 73 All images were obtained using the Aixplorer US system (SuperSonic Imag-

ine, France), frequency range 4 - 15 MHz, axial resolution 0.3 - 0.5 mm, lateral resolution 0.3 - 0.6 mm, elasticity range 0 - 300 kPa. The SWE measurements were performed using a circular region of interest (ROI) with a diameter of 3 mm. The ROI was positioned at the stiffest part of the image excluding artefacts. This procedure is equivalent to the standard imaging process of breast imaging, i.e. our aimed application, in clinical practice. All images were obtained by three observers (two radiologists with at least 5 years' experience in breast US imaging and a trained engineer). Each measurement started with minimal preload pressure, which was then increased until either the test object was damaged, the image quality was insufficient or the maximum measurement range was reached. Although the aim was to start with 0 N / 0 Pa, this was impossible in practical terms as contact was required to enable transmission of the US into the test object. Thus, the minimum pressure applied was 0.2 N / 0.4 kPa. For each setting three measurements were obtained. These measurements were averaged and evaluated using spread-sheet functions (Microsoft Excel 2013).

Results

The measurement set-up was calibrated with and without the SWE probe. Both sensors achieved good reproducibility if the device was used alone without the probe (**mean deviation from the average for sensors 1 and 2: 0.14 V and 0.40 V**). However, if the probe was attached to the pressure measurement device, the reproducibility was reduced (**mean deviation from the average for sensors 1 and 2: 0.76 V and 0.73 V**). The position of the imaging probe cable was observed to have an influence on the

reproducibility. The measurement set-up worked well on the *ex vivo* tissues and handling was similar as in the clinical imaging protocol. Figure 2 shows the correlation of the elasticity parameter mean elasticity E_{mean} for all *ex vivo* samples. The E_{mean} and E_{max} values increased approximately linearly with an increasing pre-load, whereas no clear correlation could be observed in the SD values. Figure 3 shows the SWE image in the bovine udder for the minimum (0.2 N) and an intermediate (3 N) pre-load.

Discussion

Ultrasonic SWE imaging is a hand-held imaging modality and a complete avoidance of applied pressure is impossible. Hence, the pressure applied by the ultrasound probe during clinical assessment should be considered to further standardize the diagnostic image evaluation. The demonstrated device enables monitoring of the external pressure applied in real time by an observer or clinician during SWE imaging. The **measurements are** sensitive to the weight of the probe and the cable, which was not considered when designing the device. The design could be improved in future by taking this into account. The present device was made of plastic by 3D printing. Our approach provided relatively poor physical accuracy and the material is relatively soft. **We did not observe any influence from handling the outer shell, e.g. squeezing it, on the measurements.** Better results might be achieved if a 3D printing set-up of higher quality was used or if the mechanical components of the device were made of aluminium. However, this would have increased its cost very significantly.

Our study showed that even amongst observers who apply the same imag-

ing protocol a bias in the SWE measurements may occur. Thus, real-time feedback to the observer would be helpful to standardise the imaging procedure. Definition of a pressure, which should be applied for the best clinical performance, or adjusted cut-off thresholds for benign / malignant differentiation would be of interest in the future. Additionally, monitoring the applied pressure and consequential changes in elasticity during clinical examination might also improve the benign / malignant differentiation, based on the correlation with malignancy noted in previous studies (Krouskop et al., 1998; Barr and Zhang, 2012; Syversveen et al., 2012; Sayed et al., 2014; Bernal et al., 2016). Hence, real-time measurements of the applied pressure might give clinicians a novel SWE biomarker.

Previous studies introduced different pressure application or measurement arrangements such as (Barr and Zhang, 2012; Syversveen et al., 2012; Sayed et al., 2014; Bernal et al., 2016; Bell et al., 2016, 2014; Gilbertson and Anthony, 2015). However, to the best of the authors' knowledge only the set-up introduced by Gilbertson and Anthony (2015) permits **a quantitative analysis of the applied pressure** for hand-held US imaging and could thus, be applicable to breast cancer imaging. This set-up requires a special handling and is relatively heavy (about 700 g, mass of an US probe < 100 g), **whereas the introduced set-up is much lighter (about 220 g). Although this nearly triples the weight of the transducer during the measurement, the** device introduced in this work has amongst the introduced solutions the lowest impact onto the clinical imaging protocol and has thus, the highest potential for transition into clinical routine.

A clinical trial was not possible with the device that has been described as

147 the attachment and removal process of the set-up is still too time
148 consuming. Measurements were performed only using *ex vivo* samples.
149 Inaccurate SWE values might be derived from *ex vivo* tissues due
150 to the lack of perfusion. Nevertheless, this study shows that a
151 clinical study using the introduced measurement set-up would be
152 feasible. This has potential to improve not only the benign / malignant
153 differentiation of solid breast lesions but also to improve prediction of lesion
154 behaviour and the use of personalised therapy.

155 Conclusions

156 SWE increases with applied pressure and inter-observer variations in the
157 clinical application of SWE may thus, bias the diagnostic performance of
158 SWE. Hence, real time monitoring of the applied pressure would be clinically
159 useful. The measurement device introduced here is the first step towards
160 introducing a method for examining the pressure applied during clinical
161 examinations. The results from a preliminary *ex vivo* study showing an
162 approximate linear increase in elasticity are promising. However the device
163 design should be improved to enhance its clinical applicability.

164 Acknowledgements

165 This project is part of a PhD studentship funded by SuperSonic Imagine
166 and the Engineering and Physical Science and Research Council (EPSRC,
167 EP/K020439/1).

168 References

- 169 Barr RG, Zhang Z. Effects of precompression on elasticity imaging of the
170 breast: development of a clinically useful semiquantitative method of pre-
171 compression assessment. *Journal Of Ultrasound In Medicine: Official Jour-*
172 *nal Of The American Institute Of Ultrasound In Medicine*, 2012;31:895–
173 902.
- 174 Bell MAL, Kumar S, Kuo L, Sen HT, Iordachita I, Kazanzides P. Toward
175 standardized acoustic radiation force (arf)-based ultrasound elasticity mea-
176 surements with robotic force control. *IEEE Transactions on Biomedical*
177 *Engineering*, 2016;63:1517–1524.
- 178 Bell MAL, Sen HT, Iordachita I, Kazanzides P. Force-controlled ultrasound
179 robot for consistent tissue pre-loading: Implications for acoustic radia-
180 tion force elasticity imaging. In: *Biomedical Robotics and Biomechatronics*
181 *(2014 5th IEEE RAS & EMBS International Conference on. IEEE, 2014.*
182 *pp. 259–264.*
- 183 Berg WA, Cosgrove DO, Doré CJ, Schäfer FKW, Svensson WE, Hooley RJ,
184 Ohlinger R, Mendelson EB, Balu-Maestro C, Locatelli M, Tourasse C,
185 Cavanaugh BC, Juhan V, Stavros AT, Tardivon A, Gay J, Henry J, Cohen-
186 Bacrie C. Shear-wave elastography improves the specificity of breast us:
187 the bel multinational study of 939 masses. *Radiology*, 2012;262:435–449.
- 188 Bernal M, Chammings F, Couade M, Bercoff J, Tanter M, Gennisson JL. In
189 vivo quantification of the nonlinear shear modulus in breast lesions: feasi-

190 bility study. IEEE transactions on ultrasonics, ferroelectrics, and frequency
191 control, 2016;63:101–109.

192 Gilbertson MW, Anthony BW. Force and position control system for free-
193 hand ultrasound. IEEE Transactions on Robotics, 2015;31:835–849.

194 Krouskop TA, Wheeler TM, Kallel F, Garra BS, Hall T. Elastic moduli
195 of breast and prostate tissues under compression. Ultrasonic imaging,
196 1998;20:260–274.

197 Sayed A, Layne G, Abraham J, Mukdadi OM. 3-d visualization and non-
198 linear tissue classification of breast tumors using ultrasound elastography
199 in vivo. Ultrasound in medicine & biology, 2014;40:1490–1502.

200 Syversveen T, Midtvedt K, Berstad AE, Brabrand K, Strm EH, Abildgaard
201 A. Tissue elasticity estimated by acoustic radiation force impulse quantifi-
202 cation depends on the applied transducer force: an experimental study in
203 kidney transplant patients. European Radiology, 2012;22:2130–2137.

204 Wojcinski S, Brandhorst K, Sadigh G, Hillemanns P, Degenhardt F. Acoustic
205 radiation force impulse imaging with virtual touch tissue quantification:
206 measurements of normal breast tissue and dependence on the degree of
207 pre-compression. Ultrasound in medicine & biology, 2013;39:2226–2232.

208 **Figure Captions**

209 **Figure 1:** Two shells are attached around the ultrasound probe: a) an inner
210 and b) an outer shell. Two pressure sensors are attached to the inner
211 shell. The observer presses the outer shell onto the sensors.

212 **Figure 2:** The mean elasticity E_{mean} increases with pressure in the *ex vivo*
213 samples (**correlation of 0.997, 0.981, 0.135 and 0.935 for the**
214 **chicken breast, porcine belly, boiling beef and the bovine ut-**
215 **ter**).

216 **Figure 3:** SWE image of a bovine udder with a) minimal, i.e. 0.2 N, and
217 b) intermediate, i.e. 3 N, pre-load. **The E_{mean} values were 77 kPa**
218 **and 230 kPa.**